

S O U T H W E S T R E S E A R C H I N S T I T U T E

Department of Quality Assurance
Calibration Laboratory

CERTIFICATE OF CALIBRATION

Issued to: DIV20 B57 NARASI SRIDHAR

Device No: 3239

Manufacturer: KESSLER

Model: ASTM IC 76MM

Nomenclature: THERMOMETER

Serial Number: 115847

SwRI No: NONE

Cal interval 12 Mo.

Remarks

Accuracy: MFG

Procedure: WI930-TH03

ENVIRONMENT

Temperature: 72 Humidity: 62 Location: SWRI DIV04 BLDG. 145

CONCLUSION

Tolerance/Remarks: Received into the system, introduced or reactivated

Standard Reading:	Unit Under Test:
-15 deg. C.	-15 deg. C.
0 deg. C.	0 deg. C.
75 deg. C.	75 deg. C.
140 deg. C.	140 deg. C.

Calibration was in accord with requirements of MIL-STD-45662A. Measurements are traceable to the National Institute of Standards and Technology. Inspection and test data are on file and available for inspection.

Signed

Calibration Date: 07/22/94

Cal interval: 12 Months

Record Number: 00014640

Next Calibration Due: 07/22/95

S O U T H W E S T R E S E A R C H I N S T I T U T E

Department of Quality Assurance
Calibration Laboratory

Device Serial No: 115847

Calibration Date: 07/22/94

STANDARDS

Standard No: 219 Manufacturer: AZONIX

Model: A1011

Nomenclature: RTD THERMOMETER

Serial No: T1587-2078

Cal.Due: 02/22/95

Cal.Rec.No: 00013377

Standard No: 328 Manufacturer: AZONIX

Model: A12001

Nomenclature: RTD TEMPERATURE PROBE

Serial No: 351477

Cal.Due: 03/10/95

Cal.Rec.No: 00013378

SOUTHWEST RESEARCH INSTITUTE
Department of Quality Assurance
Calibration Laboratory

CERTIFICATE OF CALIBRATION
08/14/95

Issued to: DARRELL DUNN DIV20 ,B57
Manufacturer: KESSL
Nomenclature: THERMOMETER
Serial Number: 115847
Notes:

Asset Number: 003239
Model Number: ASTM IC 76MM
SwRI/Div. I.D. #: NONE

ENVIRONMENTAL CONDITIONS

Temperature: 72.0F

Relative Humidity: 67%

CALIBRATION INFORMATION

Procedure Number: WI-9-30-TH03
Remarks:

Accuracy: MFGR SPECS
Received IN Tolerance

Calibration was in accordance with requirements of MIL-STD-45662A.
Measurements are traceable to the National Institute of Standards and Technology. Inspection and test data are on file and available for inspection.

STANDARDS USED FOR CERTIFICATION

Asset #	Serial #	Mfg	Model #	Nomenclature	Cal Date	Int.	Cal Due
000219	T1587-2078	AZONI	A1011	RTD THERMOMETER	03/06/95	12	03/06/96
000328	351477	AZONI	A12001	RTD TEMPERATURE PROBE	03/06/95	12	03/06/96

Certified by :



Certificate#: 18258

Calibration Date: 08/14/95
Interval: 12 months
Next Calibration Due: 08/14/96