

<b>TELEPHONE CONVERSATION RECORD</b>	<b>Date:</b> April 22, 2005	<b>Time:</b> 15:00
<b>Mail Control No.:</b> 136639 <b>Inspection No.:</b> N/A	<b>License No.:</b> 06-28798-01	<b>Docket No.:</b> 030-32975
<b>Person Called:</b> Alan McGinnes, Chief Executive Officer	<b>Licensee:</b> Grove Hill Medical Center, P.C.	<b>Telephone No.:</b> 203-223-0220
<b>Person Calling:</b> Steven Courtemanche/(610) 337-5075		
<b>Subject:</b> Clarification of amended license.		
<p><u>Summary:</u> I spoke with Mr. McGinnis about the licensee's request to amend the license to add a patient attenuation device. I explained that all of the sources requested could not be put on the license because the Sealed Source and Device Registration for the device mentioned is only authorized for IPL Model No. NES-8497. I also mentioned that the Authorized Users and the personnel operating the device should be trained in the use of the device before its operation. Mr. McGinnis stated he would inform his medical physicist of the conversation.</p>		
<b>Action Required/Taken:</b> Document for ADAMS		
<b>Prepared by Steven Courtemanche</b>	<b>Date:</b> April 22, 2005	