



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
801 WARRENVILLE ROAD
LISLE, ILLINOIS 60532-4351
FEB 28 2005

Stephen J. Povroznik
Radiation Safety Officer
U.S.G. Interiors, Inc.
35 Arch Street
Cloquet, MN 55720-1592

SUBJECT: TERMINATION OF YOUR NRC RADIOACTIVE MATERIALS LICENSE


Dear Mr. Povroznik:

In NRC form 314, "Certificate of Disposition of Materials" you contacted the U.S. Nuclear Regulatory Commission and indicated that you wished to terminate your NRC radioactive material license. The NRC staff has reviewed your final status survey and leak test results of five cesium-137 and one americium-241 gauging devices, we conclude that all licensable radioactive material has been removed from your facility.

Based on these conclusions, no further remediation or actions with respect to NRC regulated material is required. Your facility is suitable for unrestricted use and NRC license number 22-23428-02 is hereby terminated.

If you have any questions, please call me at (630) 829-9873.

Sincerely,


James R. Mullauer, M.H.S.
Health Physicist
Materials Licensing Branch

License No. 22-23428-02
Docket No. 030-33711

Enclosure: Amendment No. 05