

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 04/20/2005	
<b>TELEPHONE CONVERSATION RECORD</b>		Time: 2:30 PM	
Mail Control or Report No. 136501	License No. 29-30204-01	Docket No. 03033773	
Licensee/Applicant Participant(s): Morris County Cardiology Consultants Jessie Trivino- Physicist	Organization: Morris County Cardiology Consultants	Telephone No. 201-996-5720	
Person(s) Calling: Dennis Lawyer			
Subject: Questions on renewal application			
<p>Summary:</p> <p>I questioned the position of Sheila Schwartz in the organization as she had signed as the certifying officer on the application. Mr. Trivino confirmed that she would not be the appropriate person to have signed the application. I stated that an appropriate copy of Form 313 could be signed by an senior representative of the organization and faxed to 610-337-5366.</p> <p>I also questioned the statement, "If the worst deviation is more than +1-0.05 the dose calibrator.....", which is listed on page 9 of their Radiation Safety Manual. I stated that this may be a typographical error. Typically the requirement is +/- 0.05.</p> <p>Mr. Trivino said that he would get back to me on both of these issues. I gave my telephone number to him.</p>			
Action Required/Taken: Licensee has action to be taken.			
Prepared By: Dennis Lawyer		Date: 4/20/05	

*Package  
accession number ML 050750 423*