

# Mid-Delaware Imaging

J-3  
07-28657-01  
03032378

~~THIS MESSAGE IS UNCLASSIFIED EXCEPT FOR PORTIONS OF INDIVIDUALS WHOSE IDENTIFICATION IS NECESSARY TO ENFORCE FEDERAL LAWS AND TO PROTECT NATIONAL DEFENSE INFORMATION THAT IS BEING RELEASED TO THE PUBLIC UNDER THE PROVISIONS OF THE NATIONAL ARCHIVES ACT OF 1986 AND EXECUTIVE ORDER 12958.~~

~~IF YOU ARE NOT THE INTENDED RECIPIENT OF THIS MESSAGE, YOU SHOULD NOT DISSEMINATE, COPY, REPRODUCE, OR TRANSMIT THIS INFORMATION TO ANY OTHER PERSON OR ENTITY. IF YOU HAVE RECEIVED THIS COMMUNICATION IN ERROR, PLEASE NOTIFY THE IMMEDIATE SUPERVISOR OF THE PERSON WHO REPORTEDLY RECEIVED THE MESSAGE BY TELEPHONE.~~

THANK YOU.

=====

ATTN: Richard McKinley DATE: 5-3-05

FROM: Michelle Courts TIME: 1430hrs.

RE: Amendment to License

Control # 136732

MESSAGE: \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

TOTAL NUMBER OF PAGES FAXED/MAILED INCLUDING THIS PAGE: 6

WHEN YOU RECEIVE ALL PAGES AS INDICATED, PLEASE CALL Michelle

AT 302-734-9888 (FAX: 302-734-1024). THANK YOU.

710 South Queen Street  
Dover, Delaware 19904  
302-734-9888

136732  
NMSS/RGNI MATERIALS-032

MID-DELAWARE IMAGING  
NUCLEAR MEDICINE  
DAILY AREA SURVEY

WIPE TEST

SURVEY

<2kdpm/100cm<sup>2</sup>

AREA	(mR/hr)	PASS	FAIL	DECONTAMINATION
STORAGE MODULE	0.05			
HOT LAB COUNTER				
REFRIGERATOR				
L-SHIELD				
HOT SHARPS				
COLD SHARPS				
HOT LAB FLOOR				
HOT LAB TRASH				
HOT LAB DOOR				
HAND SINK				
GAMMA CAMERA				
SPECT TABLE				
WHOLE BODY TABLE				
EKG CONSOLE				
TREADMILL	not used	N/A		
INJECTION STAND				
FLOOR - INJ AREA				
TRASH - INJ AREA				
DOOR-NUC. MED. DEPT.				
TOSHIBA CONSOLE				

ATE: 11/23/04 TECH: mc  
 WIPES INSTRUMENT: atomlab 9.3  
 ACKGROUND 24 dpm  
 SURVEY INSTRUMENT: 95073  
 ACKGROUND: 0.05 mR/hr

*Handwritten signature*

ACTION LEVELS

SWIPES: 1-123. IN-111 Tc99m TI-201  
 I-131 Co-57/Ga-67  
 UNRESTRICTED: 200dpm 2,000dpm  
 RESTRICTED: 2,000dpm 20,000 dpm  
 SURVEYS:  
 UNRESTRICTED: .5mR/hr  
 RESTRICTED: 5.0mR/hr

MID DELAWARE IMAGING  
710 SOUTH QUEEN STREET  
DOVER, DE 19904  
302-734-9888

WIPE REPORT: RESULTS

DATE: Nov 23, 2004

TIME: 14:35

AREA # 1 Description: HOT LAB + Nuc. Med. Rm

Restricted, RSD trigger level at 100 % of Federal guidelines.

Isotope:	I-123	Tc-99m
Trigger in dpm:	2000	20000
Det. probability	0.99	0.99

Survey  
0.05  
mR/hr

Wipe Description	dpm	dpm
1) <u>H.L. door</u>	8	-4
2) <u>long cab. doors</u>	2	-7
3) <u>Floor 1</u>	17	-7
4) <u>Floor 2</u>	-2	-16
5) <u>Floor 3</u>	8	-2
6) <u>Floor 4</u>	-2	7
7) <u>Rt. side (Pb) Counter</u>	2	13
8) <u>long cab. counter</u>	-14	16
9) <u>Lt. side (Pb) Counter</u>	17	-2
10) <u>Corner Cabinet counter</u>	20	13
11) <u>HOT sharps</u>	-2	-10
12) <u>inside Pb cab.</u>	20	-4
13) <u>L-Shield</u>	2	16
14) <u>inj. stand</u>	-11	4
15) <u>sink</u>	8	19
16) <u>gamma camera</u>	5	13
17) <u>WB table</u>	-8	-2
18) <u>Floor - inj. area</u>	-8	4
19) <u>SPECT table</u>	8	4
20) <u>Toshiba console</u>	17	4
21) <u>H.L. trash</u>	2	4
22) <u>Storage module</u>	14	10
23) <u>Room door</u>	-5	-4
<u>inj. waste</u>		



MID DELAWARE IMAGING  
710 SOUTH QUEEN STREET  
DOVER, DE 19904  
302-734-9888

WIPE REPORT: RESULTS

DATE: Nov 23, 2004

TIME: 14:35


Wipe Description	dpm	dpm
24) <u>NM Rm. door</u>	14	-2

*.05 mR/hr*

Technologist: M. Courts

Comments: \_\_\_\_\_

HOT LAB was wiped in additional areas  
due to the fact that construction will be starting  
tomorrow on the HOT LAB to make a hallway  
to the new MRI suite.



X  
F4

- 1.) HL door
- 2.) L.C. doors (long cabinet)
- 3.) F1
- 4.) F2
- 5.) F3
- 6.) F4
- 7.) Rt. side counter (Pb)
- 8.) Long cab. counter
- 9.) L.S. counter (left side counter) (Pb)
- 10.) Cabinet counter
- 11.) HS (HOT SHAPS)
- 12.) inside Pb cabinet
- 13.) L-shield
- 14.) I.S. (inj. stand)
- 15.) Sink
- 16.) GC (gamma camera)
- 17.) WB Table
- 18.) INJ-floor
- 19.) SPECT Table
- 20.) Tosh. console
- 21.) HW (HL. waste)
- 22.) Storage module
- 23.) inj. waste
- 24.) Room door

# MID-DELAWARE IMAGING NUCLEAR MEDICINE

## SURVEY FLOOR F

