



UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION III  
2443 WARRENVILLE ROAD STE 210  
LISLE, ILLINOIS 60532-4352

FEB 28 2005

Lynn J. Patton  
Radiation Safety Officer  
10580 Wayzata Blvd., Suite 1  
Minnetonka, MN 55305

SUBJECT: TERMINATION OF YOUR NRC RADIOACTIVE MATERIALS LICENSE

Dear Mr. Patton:

By NRC Form 314 "Certificate of Disposition" (with attached documents), dated January 14, 2005, and letter dated January 18, 2005, you requested that the U.S. Nuclear Regulatory Commission terminate your NRC radioactive material license, License No. 22-26572-01. Based on information provided to us, we have confirmed that you have properly disposed of, by transfer, all your licensed material (two Troxler Model 3440 portable moisture /density gauges, each containing a nominal 8 millicurie Cs-137 sealed source and a nominal 40 millicurie Americium-241 sealed source) to another licensee (License No. 33-27581-01) authorized to possess them and you provided written confirmation of their receipt/possession of both gauges. You provided copies of leak test records that were current at the time of transfer for all four sources, confirming their non-leaking status. You also confirmed that you have never possessed any leaking sealed sources. The NRC staff has reviewed your information and has concluded that all licensable radioactive material has been removed from your facility located at 10580 Wayzata Blvd., Suite 1, Minnetonka, Minnesota and that your facility meets NRC criteria for release for unrestricted use. Accordingly, an amendment is enclosed terminating your NRC License No. 22-26572-01.

If you have any questions, please call me at (630) 829-9829.

This letter and the enclosed license are exempt from public disclosure in accordance with 10 CFR 2.390, because its disclosure to unauthorized individuals could present a security vulnerability.

Sincerely,

  
Loren J. Hueter  
Materials Licensing Branch

License No. 22-26572-01  
Docket No. 030-33512

Enclosure: Amendment No. 03