

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: May 25, 2005	
<b>TELEPHONE CONVERSATION RECORD</b>		Time: 1120	
Mail Control or Report No. 2005001	License No. 45-25339-01MD	Docket No. 03033893	
Licensee/Applicant Participant(s): Willie Regits	Organization: Cardinal Health	Telephone No. 818-7374515	
Person(s) Calling: Todd Jackson			
Subject: Clarification of Paragraph 2 in letter to Jim Dwyer dated May 24, 2005			
<p>Summary:</p> <p>I called Dr. Regits to assure that it was understood that the inspection issue Cardinal had committed to assess was the potential for overexposure of two pharmacists' extremities as a result of the recorded doses for 2004 reflecting their practice of wearing the active dosimeter of their finger rings on the outside (knuckle-side) of their hand rather than on the palm-side of the hand. Given the magnitude of the recorded doses, a correction to account for the non-conservative placement on the outside of the hand could result in a corrected dose exceeding the extremity regulatory limit of 50,000 mrem. Dr. Regits concurred that the validity and magnitude of this correction would be assessed. He noted that the text in his letter was intended to state the intention by Cardinal to try and use the field data collected to date in their study of subjects wearing multiple dosimeters on their hands to determine the magnitude of any necessary correction.</p>			
Action Required/Taken: Cardinal is assessing the inspection questions as described in their letter.			
Prepared By: Todd Jackson		Date: May 25, 2005	