

M R N J

MAGNETIC RESONANCE OF NEW JERSEY, P.A.
and Radiology Center

410 Centre Street • Nutley • New Jersey 07110 • (201) 661-2000 • Fax (201) 661-1116

Michelle R Beardsley
U.S. Nuclear Regulatory Commission Region 1
Division of Nuclear Materials Safety
475 Allendale Road
King of Prussia, Pa 19406-1415

MS-16
J1

Mail Control No. 122769
Docket No. 030-34048
February 26, 1996

Dear Ms Beardsley,

This is in reference to your letter dated February 5, 1996 requesting additional information:

1. The proposed facility is in existence with some minor work to be completed (counter tops and shelving). The gamma camera has been installed, however I confirm that no radioactive materials have been delivered, and the facility is currently not in use.
2. I confirm that no other entity other than myself controls the proposed location of use.
3. I confirm that the records of worker training will be maintained and will include the date and duration of training, the topics covered, the names(s) of the individual(s) providing training and the names of attendees.
4. I confirm that instruments will be calibrated by an instrument calibration service licensed by the NRC or Agreement State. One such service is provided by Syncor International Corporation, NRC license No 24-16617-01 MD, however other NRC or Agreement State licensed services may be used.
5. I confirm that package receipt procedures will be modified to comply with the requirements of 10 CFR 20 1906, specifically that the surfaces of a labeled package shall be monitored for radioactive contamination within 3 hours of receipt if it is received during normal working hours or not later than 3 hours from the beginning of the next working day if it is received after working hours.

FEB 28 1996

OFFICIAL RECORD COPY ML 10

M R N J

MAGNETIC RESONANCE OF NEW JERSEY, P.A.
and Radiology Center

410 Centre Street • Nutley • New Jersey 07110 • (201) 661-2000 • Fax (201) 661-1116

Michelle R Beardsley
U.S.Nuclear Regulatory Commission Region 1
Mail Control No. 122769
Docket No. 030-34048

February 26, 1996

Page 2

6. I confirm that we will have a portable detection survey instrument (Ludlum Model 14C with pancake probe) and a portable radiation measurement survey instrument (Victoreen Model 471A) as required by 10 CFR 35.220.

Thank you for your assistance, I look forward to receiving my materials license.

Sincerely,



Robert F. Traflet M.D.

dr