

Radiation Safety Officer
Saint Luke's Health System
(816) 932-6296



Fax

To: Colleen Carol Casey	From: Greg Sackett
Fax: 630-829-9782	Date: April 8, 2005
Phone:	Pages: 2
Re: Control # 314265	CC:

Urgent **For Review** **Please Comment** **Please Reply** **Please Recycle**

•Comments:

Colleen,

Let me know if you need anything else.

Greg



SAINT LUKE'S HEALTH SYSTEM

saintlukeshealthsystem.org

April 8, 2005

Materials Licensing Section
U.S. Nuclear Regulatory Commission, Region III
2443 Warrensville Road, Ste 210
Lisle, IL 60532-4352

Attn: Colleen Carol Casey

Re: License Amendment Application for License No. 24-00889-01 (Control # 314265)

In response to your fax sent April 6, 2005:

1. The Saint Luke's Radiation Safety Committee will evaluate and approve each authorized user for the Iotrex Gliasite in accordance with 35.490 or 35.940 for use of materials in 35.400.
2. A leak test via survey instrument is performed and recorded for each patient to confirm viability of the balloon catheter. Records of these leak tests will be retained for 3 years as required by 35.2067.
3. The written directive does include the treatment site/dose prescription both before and after implantation of the Iotrex solution.

If you have any further questions regarding this application, please contact me at (816) 932-6296.
Thank you for your consideration.

A handwritten signature in cursive script, appearing to read "Greg Sackett".

Greg Sackett, CHP
Radiation Safety Officer