

From: Sandra Gabriel
To: radrem@aol.com
Date: Thu, Jun 2, 2005 1:19 PM
Subject: Amendment request for St. Vincent's Medical Center, mail control 137036

Reference: St. Vincent's Medical Center
license 06-00843-03
docket 03001245
mail control 137036

To: Gerry Randall, RSO

Request is to add Adam Tazi, Ph.D. as AMP. Request stated that Dr. Tazi received an M.Sc. in Radiological Physics from Wayne State in December 2004 and that he has worked at St. Vincent's since January 3, 2005 under the direction of Vas Krithivas, Ph.D.

10 CFR 35.961(c) requires 1 year of full time training in therapeutic radiologic physics and an additional year of full time work experience under the supervision of a medical physicist at a medical institution that includes the tasks listed in 35.67, 35.632, 35.633, 35.635, 35.642, 35.643, 35.644, 35.645, and 35.652, as applicable.

Please provide:

a) copy of the transcript from Wayne State

b) statement(s) signed by authorized medical physicist(s) documenting a total of at least one year supervised full-time work experience including the tasks listed above. This experience should be in addition to the one year of full time training.

Please provide this information within 30 days. You may fax your response to my attention at 610-337-5269, referencing mail control 137036. Please leave a voicemail or e-mail message to alert me when your fax is sent.

Thank you for your help. You may e-mail me with any questions.

Sandy Gabriel
Senior Health Physicist
Medical Branch
NRC Region I
610-337-5182

Mail Envelope Properties

(429F3F8F.576 : 15 : 11102)

Subject: Amendment request for St. Vincent's Medical Center, mail control
137036

Creation Date: Thu, Jun 2, 2005 1:19 PM

From: Sandra Gabriel

Created By: SLG2@nrc.gov

Recipients

aol.com

radrem (radrem@aol.com)

Post Office**Route**

aol.com

Files

MESSAGE

Size

2624

Date & Time

Thursday, June 2, 2005 1:19 PM

Options

Expiration Date:

None

Priority:

Standard

Reply Requested:

No

Return Notification:

None

Concealed Subject:

No

Security:

Standard