



SOUTHWEST RESEARCH INSTITUTE

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-3692

COPY



Certificate #

0972-01

Certificate of Calibration

Submitted By: DIV20

Address: B57

Contact: DARRELL DUNN

Manufacturer Model: KEITHLEY 614

Description: ELECTROMETER

Serial No: 467374

Asset No: 001438

Procedure: CL-741, APR/03

Work Order: 444061748

Date Issued: Dec 2, 2004

Calibration Date: Dec 2, 2004

****Calibration Due:** Dec 2, 2005

Calibration Location: Bldg. 64

Environment: Temp. 74.0°F Hum. 38 %RH

***As Found:** IN TOLERANCE

***As Left:** IN TOLERANCE

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999 and ANSI/NCCL Z540-1-1994 which are equivalent to relevant requirements of the ISO 9000-1994 series of standards. This certificate may not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. The results of this calibration relate only to the individual instrument described above. This certificate shall not be used to claim product endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government.

Uncertainty evaluation includes the item under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor of k=2 to approximate a 95% confidence level. See Remarks or attached Calibration Report with the same Work Order number for calibration data.

*The client has sole responsibility for determination of in/out of tolerance or compliance/noncompliance. An in/out of tolerance opinion is provided for your convenience based only on the Test Instrument (TI) reading(s) and limits as reported. The reported uncertainty relates only to the results at the time of calibration and does not imply any short or long term stability of the TI.

**Calibration interval is determined by the client and does not assure the instrument will remain within tolerance until this date. Any number of factors may cause the instrument to be out of tolerance before the next calibration date.

Remarks: None

Standards Used

Asset	Manufacturer	Model	Description	Cal Due
000101	BIDDLE	72-5346-1	DECADE RESISTOR	Nov 16, 05
000182	FLUKE	5700A/EP	CALIBRATOR	Dec 23, 04
000185	GENERAL RADIO	1433G	DECADE RESISTOR	Mar 05, 05
009753	GEN RAD	1422-CB	CAPACITOR STANDARD	Oct 26, 05

Approved by: Walt Hill
Metrology Group Leader

m:\a2la1.rpt Rev date 11, May 04

Measurements by: Scott Kester
Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	444061748	Mfr.	KEITHLEY	Technician	SRK
Asset No.	001438	Model	614	Cal Date.	02-Dec-04
Serial No.	467374	Type.	Electrometer		
Remarks:					

Function/Range	Test Point	TI Reading	Difference	+/-Test Limits	+/-Uncertainty	Found/Left
Zero	mVolts 0.00000	mVolts -0.00001	mVolts -0.00001	mVolts 0.00001	mVolts 0.0000012	Results Pass
Reference	10.0000					
DCV	Volts	Volts	Volts	Volts	Volts	
0.2 V	0.19000 0.10000	0.19000 0.10000	0.00000 0.00000	0.00017 0.00010	0.000022 0.000022	Pass Pass
2 Volt	1.9000 1.0000	1.8998 1.0000	-0.0002 0.0000	0.0016 0.0009	0.00012 0.00012	Pass Pass
20 Volt	19.000 10.000	18.997 9.999	-0.003 -0.001	0.016 0.009	0.0012 0.0012	Pass Pass
DC Amps	uAmps	uAmps	uAmps	uAmps	uAmps	
20 uAmp	19.00 nAmps	18.99 nAmps	-0.01 nAmps	0.08 nAmps	0.014 nAmps	Pass
200 nAmp	190.0 pAmps	190.1 pAmps	0.1 pAmps	1.1 pAmps	0.12 pAmps	Pass
2000 pAmp	1900	1894	-6	29	0.12	Pass
Resistance	kOhm	kOhm	kOhm	kOhm	kOhm	
20 kOhm	19.00	18.99	-0.01	0.11	0.012	Pass
200 kOhm	190.0	189.8	-0.2	1.1	0.12	Pass
20 MOhm	10.00 MOhm	9.98 MOhm	-0.02 MOhm	0.10 MOhm	0.012 MOhm	Pass
20 GOhm	10.00 GOhm	10.05 GOhm	0.05 GOhm	0.22 GOhm	0.014 GOhm	Pass
Charge Calibration						
2 nC	1.0000	0.9987	-0.0013	0.0500	0.00012	Pass

END OF REPORT