

U.S. NUCLEAR REGULATORY COMMISSION		Conversation Date: 06/15/2005	
<b>TELEPHONE CONVERSATION RECORD</b>		Time: 1540 hours	
Mail Control No.:	License No.:	Docket No.:	
137035	37-06621-01	03003074	
Licensee/Applicant Participant(s):	Organization:	Telephone No.:	
Keith Ostrom Consultant	Ohio Valley General Hospital	216-663-7000	
Person(s) Calling: Shirley Xu, RI			
Subject: renewal			
<p>Summary:</p> <p>Following issues were discussed with Mr. Ostrom:</p> <ol style="list-style-type: none"> <li>1. Clarify the total activities for 35.300 and 35.500 uses.</li> <li>2. Asked him to confirm that they have developed and will implement and maintain written survey meter calibration procedures in accordance with the requirements in 10 CFR 20.1501 and that meet the requirements of 10 CFR 35.61.</li> </ol>			
Action Required/Taken: await for response			
Prepared By: S Xu		Date: 06/15/2005	