



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

June 21, 2005

Docket No. 03020775
Control No. 136951

License No. 19-23344-01

Major Daniel S. Hamilton
Radiation Safety Officer
Uniformed Services University of the Health Sciences
4301 Jones Bridge Road
Bethesda, MD 20814-4799

**SUBJECT: UNIFORMED SERVICES UNIVERSITY OF THE HEALTH SCIENCES,
REQUEST FOR ADDITIONAL INFORMATION CONCERNING APPLICATION
FOR AMENDMENT TO LICENSE, CONTROL NO. 136951**

Dear Major Hamilton:

This is to confirm Dennis Lawyer's telephone conversation on May 16, 2005 with you in which we discussed the information we need to continue review of your application dated April 25, 2005. The items discussed are specified below.

1. The Decommissioning Plan Cost Estimate of 2005 submitted with the renewal is not consistent with the March 4, 1992, Decommissioning Funding Plan. This plan must include an actual estimate of the costs for decommissioning your facility. The appropriate level of detail for the cost estimate is discussed in Appendix A.3 to NUREG-1757 Volume 3, "Consolidated NMSS Decommissioning Guidance". Please update the Decommissioning Plan with appropriate supporting documents.
2. The Statement of Intent submitted with the renewal did not contain the title of the person signing or the date of signature. Please resubmit the Statement of Intent indicating the title of the person signing and the document dated.
3. Describe the surveys you will require and the criteria you will use for release of facilities and equipment for unrestricted use. Confirm that facilities and equipment will not be released until the results of surveys are reviewed and approved by the Radiation Safety Officer.

Current NRC regulations and guidance are available at the NRC web site at <http://www.nrc.gov/materials/miau/mat-toolkits.html> and <http://www.nrc.gov/who-we-are/governing-laws.html> or by contacting the Government Printing Office (GPO) toll-free at 1-888-293-6498. The GPO is open from 7:00 a.m. to 9:00 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 136951. If you have any technical questions

D. Hamilton

2

regarding this deficiency letter, please call me at (610) 337-5308 or Dennis Lawyer at 610-337-5366.

If we do not receive a reply from you within 30 calendar days from the date of this letter, we will assume that you do not wish to pursue your application.

Sincerely,

Original signed by Todd J. Jackson, CHP

Todd J. Jackson, CHP
Senior Health Physicist
Commercial and R&D Branch
Division of Nuclear Materials Safety

DOCUMENT NAME: E:\Filenet\ML051720385.wpd

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI	N	DNMS/RI			
NAME	DLawyer/DRL1		TJackson/TJJ					
DATE	06/21/2005		6/21/2005					

OFFICIAL RECORD COPY