

Lilly del Caribe, Inc.
P.O. Box 1198
Carolina, PR 00986-1198

(787) 257-5555

June 17, 2005

NMSB3

David J. Collins
Sr. Health Physicist
U.S. Nuclear Regulatory Commission
Division of Nuclear Safety
Region II
Atlanta, GA 30303-8931

03032597

Dear Mr. Collins:

Pursuant to US NRC materials license no. 52-25181-01, the attached form NRC 314 (original), Thermo Electron Corporation "Acknowledgement of Receipt of Radioactive Material" (copy), and the recent leak test results (copy) formally document the decommissioning of the nuclear level detectors (sealed sources).

With this communication, iPR Pharmaceuticals and its former Radiation Safety Officer conclude all activities related to the referred license, reason for which we request its termination.

If you have any question or concern on this subject, you may contact me at 787-866-1290 ext. 3404.

Sincerely,

Hiram Pabón
RSO / Sr. Safety Engineer

Encl.

c: Teresa Colón
Frank Cortés

CERTIFICATE OF DISPOSITION OF MATERIALS

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Records and FOIA/Privacy Services Branch (T-5 F52), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to infocollects@nrc.gov and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0028) Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

LICENSEE NAME AND ADDRESS

IPR Pharmaceuticals, Inc.
P.O. Box 10000
Guayama, Puerto Rico 00785

LICENSE NUMBER

52-25181-01

DOCKET NUMBER

030-32597

LICENSE EXPIRATION DATE

06/30/2012

A. LICENSE STATUS (Check the appropriate box)

This license has expired. This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

- 1. No radioactive materials have ever been procured or possessed by the licensee under this license.
- 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner:
 - a. Transfer of radioactive materials to the licensee listed below:
Thermo Electron
2555 N. IH-35 Round Rock, Texas 78664
 - b. Disposal of radioactive materials:
 - 1. Directly by the licensee:
 - 2. By licensed disposal site:
 - 3. By waste contractor:
 - c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

C. SURVEYS PERFORMED AND REPORTED

- 1. A radiation survey was conducted by the licensee. The survey confirms:
 - a. the absence of licensed radioactive materials
 - b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
- 2. A copy of the radiation survey results:
 - a. is attached; or b. is not attached (Provide explanation); or c. was forwarded to NRC on: _____ Date _____
- 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
 - a. The results of the latest leak test are attached; and/or
 - b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME Hiram Pabón	TITLE Safety Engineer II (Radiation Safety Officer)	TELEPHONE (Include Area Code) (787) 866-1290	E-MAIL ADDRESS hiram.ppabon@astraz
---------------------	--	---	---------------------------------------

Mail all future correspondence regarding this license to:
Juan L. Hernandez Safety & Occupational Health Manager IPR Pharmaceuticals, Inc. P.O. Box 1624 Canovanas, Puerto Rico 00729-1624

C. CERTIFYING OFFICIAL I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE Frank Cortés - Chemical Operations Director	SIGNATURE <i>Frank Cortés</i>	DATE April 28, 2005
---	----------------------------------	------------------------

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

137320

ACKNOWLEDGMENT OF RECEIPT OF RADIOACTIVE MATERIAL

May 25, 2005

Hiram Pabon
IPR Pharmaceuticals, Inc.
Guayama Plant
Road 53, KM. 82, Jobos Exit
Guayama, PR 00784

RMA Number 30198

Attention Hiram Pabon:

This is to certify that Thermo Electron has received and accepted ownership of the radioactive material described below pursuant to applicable regulations and as authorized by our Texas Radioactive Material License L03524.

Manufacturer	Model	Serial	Isotope	Source	Activity Units	Assay
KAY-RAY/SENSALL	7062BP	28209	Cs-137	GH-1625	5 mCi	10/1/1989
KAY-RAY/SENSALL	7062BP	28210	Cs-137	GH-1626	5 mCi	10/1/1989
KAY-RAY/SENSALL	7063P	28206	Cs-137	GH-1772	200 mCi	3/1/1990
KAY-RAY/SENSALL	7063P	28207	Cs-137	GH-1769	200 mCi	3/1/1991
KAY-RAY/SENSALL	7063P	28208	Cs-137	GH-1783	200 mCi	3/1/1991

This receipt should be retained in your files as a permanent record showing the disposition of this radioactive material. If you are not the Radiation Safety Officer or responsible for maintaining regulatory records for radioactive material, please forward this letter to the appropriate person.

If you have any questions or require additional assistance, please contact us at 713-272-5389.

Sincerely,
Thermo Electron



Angelica Guidry
Services Support Specialist

April 4, 2005

Hiram Pabon
IPR PHARMACEUTICALS
Rd #53, Km 82 Jobos Exit
P.O. Box 10000; 00785
Guayama, PR 00784

Dear Hiram Pabon,

I am enclosing the results from your recent leak tests. Please keep these certificates in a permanent file for inspection by the regulatory agency.

Thank you for using Thermo Electron's leak test service. If we can be of further assistance, please call us at 800-736-0801 or send email to leaktest@thermoMT.com.

Sincerely,


Martha Hernandez

LEAK TEST CERTIFICATION

This is to certify that the product identified below was tested for removable contamination in accordance with Texas Radioactive Material License Number L03524, Condition 19.

Customer: 989501
IPR PHARMACEUTICALS
Rd #53, Km 82 Jobos Exit
P.O. Box 10000; 00785
Guayama, PR 00784

Device ID: 

Product: KAY-RAY/SENSALL

Model: 7062BP Serial No 28210

Isotope: Cs-137 Activity: 5 mCi

Source Serial: N/A10519


Location: 21-07X01(TOP)

Test Type: Lab Counting

Result: Positive Negative < 2.5E-5 uCi

Analyzed on: April 04, 2005

This certificate should be maintained as a permanent record of the leak test of this product.


Martha Hernandez

LEAK TEST CERTIFICATION

This is to certify that the product identified below was tested for removable contamination in accordance with Texas Radioactive Material License Number L03524, Condition 19.

Customer: 989501
IPR PHARMACEUTICALS
Rd #53, Km 82 Jobos Exit
P.O. Box 10000; 00785
Guayama, PR 00784

Device ID: 

Product: KAY-RAY/SENSALL

Model: 7062BP Serial No 28209

Isotope: Cs-137 Activity: 5 mCi

Source Serial: N/A10518

Location: 21-07X01(BOTTOM)

Test Type: Lab Counting

Result: Positive Negative < 2.5E-5 uCi

Analyzed on: April 04, 2005

This certificate should be maintained as a permanent record of the leak test of this product.


Martha Hernandez

LEAK TEST CERTIFICATION

This is to certify that the product identified below was tested for removable contamination in accordance with Texas Radioactive Material License Number L03524, Condition 19.

Customer: 989501
IPR PHARMACEUTICALS
Rd #53, Km 82 Jobos Exit
P.O. Box 10000; 00785
Guayama, PR 00784

Device ID: 

Product: KAY-RAY/SENSALL

Model: 7063P Serial No 28207

Isotope: Cs-137 Activity: 200 mCi

Source Serial: N/A10522


Location: 23-03R01

Test Type: Lab Counting

Result: Positive Negative < 2.5E-5 uCi

Analyzed on: April 04, 2005

This certificate should be maintained as a permanent record of the leak test of this product.


Martha Hernandez

LEAK TEST CERTIFICATION

This is to certify that the product identified below was tested for removable contamination in accordance with Texas Radioactive Material License Number L03524, Condition 19.

Customer: 989501
IPR PHARMACEUTICALS
Rd #53, Km 82 Jobos Exit
P.O. Box 10000; 00785
Guayama, PR 00784

Device ID: 

Product: KAY-RAY/SENSALL

Model: 7063P Serial No 28206

Isotope: Cs-137 Activity: 200 mCi

Source Serial: N/A10521

Location: 23-01R01

Test Type: Lab Counting

Result: Positive Negative < 2.5E-5 uCi

Analyzed on: April 04, 2005

This certificate should be maintained as a permanent record of the leak test of this product.


Martha Hernandez

LEAK TEST CERTIFICATION

This is to certify that the product identified below was tested for removable contamination in accordance with Texas Radioactive Material License Number L03524, Condition 19.

Customer: 989501
IPR PHARMACEUTICALS
Rd #53, Km 82 Jobos Exit
P.O. Box 10000; 00785
Guayama, PR 00784

Device ID: 

Product: KAY-RAY/SENSALL

Model: 7063P

Serial No 28208

Isotope: Cs-137

Activity: 200 mCi

Source Serial: N/A10520


Location: 23-08R01

Test Type: Lab Counting

Result: Positive Negative < 2.5E-5 uCi

Analyzed on: April 04, 2005

This certificate should be maintained as a permanent record of the leak test of this product.


Martha Hernandez

137320

From
 06 23 05 Sender's FedEx Account Number 1474-4689-6

Sender's Name
 Arlene Perez Phone 787 866-1290

Company
 THE PHARMACEUTICAL CO

Address
 P.O. Box 14300 P.O. Box 14300
 Columbia State/Province FL

City
 ZIP/Postal Code 33033

Your Internal Billing Reference Information

To
 Recipient's Name DAVID J COLLINS Phone

Company
 US NUCLEAR REGULATORY COMMISSION DIVISION OF NUCLEAR SAFETY

Address
 REGION II Dept/Floor

City
 ATLANTA State/Province GA

Country
 USA ZIP/Postal Code 30303-8931

Recipient's I.D. number for Customs purposes (e.g. GST, REC/VAT or as locally required)

For HOLD at FedEx Location check here For Saturday Delivery check here

Shipment Information

FedEx Tracking Number **8068 8813 3895** Form I.D. No. **0400**

5 Broker Selection Broker's Name FedEx International Broker Selection

City/State/ZIP/Postal Code
FedEx
 emp: 228120 27JUN05

6 S
 FedEx TRK# **8068 8813 3895** FORM 0400
 (Post)

7 Pac
 FedEx (FedEx International) **30303 -GA-US**

8 Spec
Does this shipment require special handling?
 No Yes



INTL PRIORITY

TUE
 Deliver By:
28JUN05
 A1

9a Payment - TRANSPORTATION CHARGES PAID BY:
 Sender Recipient Third Party Credit Card Cash/Check
 (Enter: FedEx Account No. or Credit Card No. below) Obtain Recipient/Third Party FedEx Account No.

FedEx Account No. **1474-4689-6**
 Credit Card No. Not all options available to all destinations. Exp. Date

9b Payment - DUTIES AND TAXES PAID BY:
 Sender Recipient Third Party Cash/Check
 FedEx Account No. Total (Excludes Customs charges) \$

10 Required Signature
 By giving us your shipment, you agree to the conditions on the back of this Non-Negotiable Air Waybill. Certain international treaties, including the Warsaw Convention, may apply to this shipment and limit our liability for damage, loss or delay, as described in the Conditions of Contract.
 WARNING: These commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. Law prohibited.
 Sender's Signature: _____ Date Executed: _____
 Received above shipment in good order and condition. We agree to pay all charges including Customs duties and taxes as applicable and to the Conditions of Contract as stated on the reverse side of the consignor copy.
 Recipient's Signature: _____



This is to acknowledge the receipt of your letter/application dated 6/17/2005 / NAC 314 / 4/28/2005, and to inform you that the initial processing which includes an administrative review has been performed.

TEAM 52-25181-01 There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 137320.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.