



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

July 25, 2005

Docket No. 03032370
Control No. 137066

License No. 52-25154-01

Jorge Hernández, M.D.
2225 Edificio Parra, Suite 802
Ponce By-Pass
Ponce, PR 00717-1322

SUBJECT: VOIDANCE OF APPLICATION FOR LICENSE TERMINATION, CONTROL NO.
137066

Dear Dr. Hernández:

This concerns the subject application for termination of your material license dated May 12, 2005. As discussed with your assistant during a telephone conversation on July 20, 2005, we are not able to grant your request at this time. Specifically, Dr. Quiñones Ferrer must request and obtain an amendment to his license to authorize possession of the make, model and activity of your eye applicator before you can transfer your applicator to him. After that, you may request termination by certifying that the transfer has taken place and providing a copy of a leak test report that is less than six months old at the time of the transfer which indicates the applicator is free from radioactive contamination.

As a result, we have voided your application. This action is taken without prejudice to the resubmission of your request.

Sincerely,

Original signed by Pamela J. Henderson for

Héctor Bermúdez
Senior Health Physicist
Medical Branch
Division of Nuclear Materials Safety

DOCUMENT NAME: E:\Filenet\ML052070055.wpd

SISP Review Complete: (Reviewer's Initials) PJH

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI		DNMS/RI			
NAME	HBermúdez/PJH for							
DATE	7/25/05							

OFFICIAL RECORD COPY