



# SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-4834



Certificate #

0972-01

## Certificate of Calibration

**Submitted By:** DIV20

**Address:** B57

**Contact:** BRADLEY WERLING

**Manufacturer Model:** BARNSTEAD INTL. E896X5

**Description:** RESISTOR

**Serial No:** 896020321720

**Asset No:** 009582

**Procedure:** RESISTORS JAN/04

**Work Order:** 303064626

**Date Issued:** Jun 13, 2005

**Calibration Date:** Jun 13, 2005

**\*Calibration Due:** Jun 13, 2006

**Calibration Location:** Bldg. 64

**Environment:** Temp. 72.0°F Hum. 42 %RH

**\*\*Data Type:** FOUND-LEFT

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

\*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. \*\*Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

**Remarks:** None

### Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
007001	2823A21362	HEWLETT-PACKARD	3458A/OPT-002	MULTIMETER	Feb 17, 06

Approved by: Walt Hill  
Manager

m:\a2la1.rpt Rev date March 6, 2005

Measurements by: Curtis Laurence  
Metrology Technician

Page 1 of 1

Southwest Research Institute  
Calibration Laboratory  
Calibration Report

Work Order:	303064626	Mfr.	Barnstead Int.	Technician	WCL
Asset No.	009582	Model	E896X5		
Serial No.	896020321720	Type.	Resistor	Cal Date.	13-Jun-05
Remarks:					
Tested to +/- 1 %					

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
	MOhm	MOhm	MOhm	MOhm	MOhm	Result
1.5 Mohm	1.5000	1.5049	0.0049	0.0150	0.0053	Pass
	Ohm	Ohm	Ohm	Ohm	Ohm	
552.95 Ohm	552.95	553.337	0.387	5.53	0.0085	Pass
END OF REPORT						