



# SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-4834



Certificate #

0972-01

## Certificate of Calibration

**Submitted By:** DIV20

**Address:** B57

**Contact:** JIM PRIKRYL

**Manufacturer Model:** OHAUS TS 400D

**Description:** BALANCE

**Serial No:** 2883

**Asset No:** 002345

**Procedure:** BALANCES & SCALES, DEC/04

**Work Order:** 303064982

**Date Issued:** Jul 12, 2005

**Calibration Date:** Jul 12, 2005

**\*Calibration Due:** Jan 12, 2006

**Calibration Location:** B57

**Environment:** Temp. 69.0°F Hum. 69 %RH

**\*\*Data Type:** FOUND-LEFT

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.


\*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. \*\*Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.


**Remarks:** None

### Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
007290	T4830007	VAISALA	HM34F	HUMIDITY/ TEMPERATURE METER	Apr 11, 06
010442	A007	RICE LAKE	10MG-100G	WEIGHT SET, CLASS 1	Apr 26, 06
011236	21R2	RICE LAKE	200G	WEIGHT, CLASS 1	Sep 14, 06
011237	21R3	RICE LAKE	200G	WEIGHT, CLASS 1	Sep 14, 06
005566	B0113L004	MERIAN	A0030P	PRESSURE GAUGE, ABSOLUTE	Oct 13, 05

Approved by:   
Manager

m:\a2la1.rpt Rev date March 6, 2005

  
Measurements by: Scott Kester  
Metrology Technician

Southwest Research Institute  
Calibration Laboratory  
Measurement Report

Work Order:	303064982	Manufacturer:	Ohaus	Technician:	SRK
Asset Number:	002345	Model:	TS 400D		
Serial Number:	2883	Type:	Balance	Cal Date:	12-Jul-05
Remarks:	Manufacturer does not provide corner load specifications. Corner load readings are without pass or fail indications.				
Ambient Conditions	69 ° F		69 % RH		14.29 PSIA

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
Corner Load	grams	grams	grams	grams	grams	Result
Reference	200.00	200.01				
Front	200.01	200.01	0.00		0.0075	
Rear	200.01	200.01	0.00		0.0075	
Left	200.01	200.01	0.00		0.0075	
Right	200.01	200.01	0.00		0.0075	
Repeatability						
1	200.00	200.01				
2	200.00	200.01				
3	200.00	200.01				
4	200.00	200.01				
5	200.00	200.01				
6	200.00	200.01				
7	200.00	200.01				
8	200.00	200.01				
9	200.00	200.01				
10	200.00	200.01				
Std Deviation		0.000		0.020		Pass
Linearity	0.00	0.00	0.00	0.02	0.0075	Pass
	40.00	40.00	0.00	0.02	0.0075	Pass
	80.00	80.00	0.00	0.02	0.0075	Pass
	120.00	120.00	0.00	0.02	0.0075	Pass
	160.00	160.00	0.00	0.02	0.0075	Pass
	200.00	200.00	0.00	0.02	0.0075	Pass
	240.00	240.00	0.00	0.02	0.0075	Pass
	280.00	280.00	0.00	0.02	0.0075	Pass
	320.00	320.00	0.00	0.02	0.0075	Pass
	360.00	360.00	0.00	0.02	0.0075	Pass
	400.00	400.00	0.00	0.02	0.0075	Pass

End of Report