



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

August 4, 2005

Docket No. 03035993
Control No. 137282

License No. 29-30730-01

John Galdi, Sr.
Administrator
Superior Medical Diagnostics II, LLC
235 Franklin Avenue
Nutley, NJ 07110

SUBJECT: SUPERIOR MEDICAL DIAGNOSTICS II, LLC, REQUEST FOR ADDITIONAL INFORMATION CONCERNING APPLICATION FOR AMENDMENT TO LICENSE, CONTROL NO. 137282

Dear Mr. Galdi:

This is in reference to your letter dated June 10, 2005, requesting to amend Nuclear Regulatory Commission License No. 29-30730-01. In order to continue our review, we need the following additional information:

1. In your letter dated June 10, 2005, you request that the license be amended to include Orestes Sanchez, M.D. on the list of authorized users. Please confirm that Dr. Sanchez authorization is to include licensed material covered by the regulations in 10 CFR 35.500, i.e., patient attenuation devices. Also, confirm that Dr. Sanchez will receive training in the use of the patient attenuation device prior to his supervision of licensed activities.
2. Confirm that all of the physicians listed as authorized users have received training as described in 10 CFR 35.590.
3. Please note the changes in your NRC license have been made to confirm more closely with current NRC licensing policy. Items 6.A. and B. are covered under 10 CFR 35.65 and Item 6.C. has been changed to reflect that patient attenuation devices are covered under 10 CFR 35.500. Since the authorization for Items 6.A. and 6.B. are covered generically under 10 CFR 35.65, Charles Giomuso has been removed from the authorized user list with no effect on your program.

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Medical, industrial, and academic uses of nuclear material**; then **toolkit index page**. Or you may obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-888-293-6498. The GPO is open from 7:00 a.m. to 9:00 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 137282. If you have any technical questions regarding this deficiency letter, please call me at (610) 337-5075.

J. Galdi
Superior Medical Diagnostics II, LLC

2

If we do not receive a reply from you within 30 calendar days from the date of this letter, we will assume that you do not wish to pursue your application.

Sincerely,

Original signed by Steven Courtemanche

Steven Courtemanche
Health Physicist
Commercial and R&D Branch
Division of Nuclear Materials Safety

cc:
Charles A. Giomuso, Radiation Safety Officer

DOCUMENT NAME: E:\Filenet\ML052210087.wpd

SISP Review Complete: SRC

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	<input checked="" type="checkbox"/> N	DNMS/RI	<input type="checkbox"/>	DNMS/RI	<input type="checkbox"/>		
NAME	SCourtemanche/src							
DATE	08/04/2005							

OFFICIAL RECORD COPY