

FRIEDA SILVA DE ROUDAN, M.D.



Frieda O. SILVA DE ROUDAN, M.D.  
Plaza Nuclear Imaging  
La Torre de Plaza  
Suite 401  
Hato Rey, PR 00918

August 10, 2005

US NRC RI  
DNMS  
475 Allendale Road  
King of Prussia, PA 19406

MS 16  
J-9

RE: Additional information for control # 137103.

Dear Mr. Sir or Madam:

03033894

Please find enclosed the answers to your letter dated July 8, 2005 regarding an outside consultant for NRC License 52-25340-01.

- a. David Rhoe is an outside consultant.
- b. David Rhoe will have complete authority as the RSO to correct and implement corrective actions.
- c. The RSO will report directly to upper management. If needed, management will provide the necessary funds to maintain the program.
- d. David Rhoe provides consulting to several other NRC and Agreement state licenses. Due to the fact that all are considerably small programs, little time is required on site. David Rhoe will visit the facility once each month to review the records of the program and, as needed, will provide additional time to maintain the program within compliance. David Rhoe has been performing these RSO duties since July 2002.
- e. The nuclear medicine technologist will serve as a contact point during the RSO's absence.
- f. The consultant-RSO will be available seven days a week, 24 hours a day to answer questions or respond to an emergency.

If you need any further information, please contact me at (787) 754-0715.

Sincerely,

Frieda Silva, MD

137103

NMCG/RGNI MATERIALS-002

Nuclear Medicine Quarterly Inspection

Facility: Plaza Nuclear

Surveyor: David Rhoe

Date: 8/1/05

Records	Compliance (Y N)	Comments
1. License and regulations	Y	
2. Posting requirements	Y	
3. Radiation Safety Committee	NA	
4. Facilities and engineering controls	Y	
5. Dose calibrator	Y	Need PSD signature
6. Generators	NA	
7. Syringe shields and labels	Y	
8. Vial shields	Y	
9. Personnel	Y	
10. Sealed sources	Y	Need PSD Signature
11. Laboratory Surveys	Y	
12. Radiation detection equipment	Y	
13. Ambient dose rates	Y	
14. Radioactive waste	Y	
15. Transportation	NA	
16. Incoming/out going packages	Y	
17. Patient log	Y	
18. Patient release	Y	
19. Therapies	—	
20. Bioassays	—	
21. Bulletins and etc.	—	

Seen July 05

July

Recommendations

New Co-57 source will be required in Feb 2006

Other items

1. Training		1-05
2. Co-57 source		59.8 uCi
3. Fume hood flow		
4. Dose calibrator accuracy		1-05
5. Survey meter	Make: Mod: 2241-2 Sn	1-05
6. Survey meter	Make: Mod: 14C Sn	2-05
7. Scaler	Make: Mod: Sn	1-05
8. DWTC	Make: Mod: Sn	
9. Thyroid uptake	Make: Mod: Sn	1-05
10.		