

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: <i>Creagh E. Milford, DO. Waterford, MI</i>		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III 2443 Warrenville Road Suite 210 Lisle, Illinois 60532-4351	
3. DOCKET NUMBER(S) <i>030-30281</i>		4. LICENSEE NUMBER(S) <i>21-25851-01</i>	
		5. DATE(S) OF INSPECTION <i>Aug 2 and 4, 2005</i>	

LICENSEE:
The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection

1. Based on the inspection findings, no violations were identified.

2. Previous violation(s) closed.

3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Deborah A. Piskura	<i>D. Piskura</i>	8/4/05

Docket File Information
SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION

1. LICENSEE Creagh E. Milford, D.O. REPORT NUMBER(S) 2005-001 and -002		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road, Suite 210 Lisle, IL 60532	
3. DOCKET NUMBER(S) 030-30281		4. LICENSE NUMBER(S) 21-25851-01	5. DATE(S) OF INSPECTION Aug. 2&4, 2005
6. INSPECTION PROCEDURES USED 87130		7. INSPECTION FOCUS AREAS 03.01, 03.02, 03.03, 03.04, 03.05, 03.06, 03.07, and 03.08	
SUPPLEMENTAL INSPECTION INFORMATION			
1. PROGRAM CODE(S) 02201	2. PRIORITY G 5	3. LICENSEE CONTACT Creagh E. Milford D.O., RSO	4. TELEPHONE NUMBER 248.666.5200
<input checked="" type="checkbox"/> Main Office Inspection	Next Inspection Date: 08/2010		
<input checked="" type="checkbox"/> Field	5625 Water Tower Place, Clarkston, MI		
<input type="checkbox"/> Temporary Job Site			

PROGRAM SCOPE

This licensee was a private practice physicians office and authorized to use licensed material permitted by Sections 35.100 and 35.200 at two clinic locations. The nuclear medicine department was staffed with three full-time technologists who rotate between both clinics and performed approximately 150-200 diagnostic nuclear cardiology procedures per month. The licensee received unit doses from a licensed radiopharmacy. The clinic retained the services of a consulting physicist who audited the radiation safety program on a quarterly basis.

This inspection consisted of interviews with licensee personnel, a review of select records, tour of the nuclear medicine department, and independent measurements. The inspector observed licensee personnel prepare, assay and administer several unit dosages of Tc-99m Cardiolite for a cardiac resting and stress testing procedures.