



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE ROAD STE 210
LISLE, ILLINOIS 60532-4352

AUG 08 2005

Cheryl M. Culver Schultz
Radiation Safety Officer
William Beaumont Hospital
Division of Cardiovascular Services
3601 West 13 Mile Road
Royal Oak, MI 48073-6769

SUBJECT: TERMINATION OF YOUR NRC RADIOACTIVE MATERIALS LICENSE

Dear Ms. Schultz:

In a letter dated May 26, 2005, received June 23, 2005, you contacted the U.S. Nuclear Regulatory Commission and indicated that you wished to terminate your special nuclear materials license no. SNM-1366.

The NRC staff has reviewed your final status surveys, consisting of confirmation that all licensed pacemakers have been explanted and returned to authorized vendors for disposal. Based on its review, the staff has concluded that all licensable radioactive material has been removed from your facilities and residual radioactive material attributable to licensed activities does not exceed current NRC criteria.

Based on these conclusions no further remediation or actions with respect to NRC regulated material is required. Your facility is suitable for unrestricted use and NRC license number SNM-1366 is hereby terminated with the enclosed Amendment No. 11.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and the enclosed license termination document will be available electronically in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

If you have any questions or comments regarding our handling of your license in this situation, please call me at (630) 829-9841.

Sincerely,

A handwritten signature in cursive script that reads "Colleen Carol Casey".

Colleen Carol Casey
Materials Licensing Branch

License No. SNM-1366
Docket No. 070-01398

Enclosure: Amendment No. 11