

KNIGHT CONSULTING ENGINEERS, INC.

P.O. Box 29

WILLISTON, VERMONT 05495

Tel: 802-879-6343 Fax: 802-879-6376

STEPHEN C KNIGHT, JR. P.E.
MARTIN W HAIN, P.E.
ERIC H. GODDARD, P.E.
ROBERT A. LOVGREN, P.E.
BARBARA J. EVANS, P.E.

CIVIL ENGINEERING
STRUCTURAL ENGINEERING
GEOTECHNICAL ENGINEERING
HYDRAULIC ENGINEERING
CONSTRUCTION INSPECTION
CONSTRUCTION TESTING
SURVEYING

MS16

J-1

August 18, 2005

Michelle Beardsley
U.S. Nuclear Regulatory Commission, Region I

VIA FACSIMILE: 1-610-337-5393

Re: NRC License No. 44-19115-02

03022089

Subject: Control #137165, Renewal Application Clarifications


Dear Ms. Beardsley

Confirming our recent conversation regarding our responses to item 10 of the application, I offer the following clarification:

- Leak Tests:
 - We perform leak testing at 6 month intervals using test kits provided by an organization authorized by the NRC. Leak test analysis is conducted by the supplier of the test kits.
- Non-Routine Maintenance:
 - We do not perform non routine maintenance or repair operations requiring removal of the source or source rod from the gauge. All work of this nature is done by an organization authorized by the NRC to perform such work

Please contact the undersigned if you have any comments or questions regarding this matter.

Yours truly



Robert A. Lovgren, P.E., RSO
KNIGHT CONSULTING ENGINEERS, INC.

RAL:nmv

\$ RAL\licensing\tr2.doc

137165

NMSS/RGNI MATERIALS-002