

U.S. NUCLEAR REGULATORY COMMISSION		Date: 8/23/05
TELEPHONE CONVERSATION RECORD		Time: 3:15 p.m.
Mail Control 137397 or Report No(s).	License No(s). 29-16145-01	Docket No(s). 03010491
Name of Licensee: Robert Wood Johnson University Hospital at Hamilton		
Name of Participant(s): Janet Bryant, physics consultant		
Telephone No. 908-788-9440, ext. 40		
Subject: Additional information needed for license amendment request <small>(NOTE: This will be used as the Documents Title in ADAMS)</small>		
Summary: I left a voicemail message for Janet requesting the following: <ul style="list-style-type: none"> a) Provide room number and diagram for area where I-125 sources will be stored. b) What is method for securing I-125 sources in storage area? c) What is sensitivity of detection instruments to be used to locate stray I-125 sources? We suggest licensee consider use of a thin crystal NaI probe. d) For Dr. Kagan, provide documentation that his 500 hours in brachytherapy training included the tasks listed in 35.490(b)(1)(ii). <p>On 8-31-05, Janet left a voicemail message for me stating that she received my message and forwarded it to physicist Gary Jones. On 9-1-05, Gary called and said he received the forwarded message and is attempting to obtain this information.</p>		
Action Required: Await response from licensee		
Document Availability: <input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date		
Prepared & SISP Review Completed By: / RA / Sandy Gabriel Date: 9-1-05		