

| | | |
|--|--|--|
| U.S. NUCLEAR REGULATORY COMMISSION | | Date: 8/23/05 |
| TELEPHONE CONVERSATION RECORD | | Time: 3:15 p.m. |
| Mail Control or Report No(s). | 137397 | License No(s). 29-16145-01 Docket No(s). 03010491 |
| Name of Licensee: | Robert Wood Johnson University Hospital at Hamilton | |
| Name of Participant(s): | Janet Bryant, physics consultant | |
| Telephone No. | 908-788-9440, ext. 40 | |
| Subject: <small>(NOTE: This will be used as the Documents Title in ADAMS)</small> | Additional information needed for license amendment request | |
| Summary: | <p>I left a voicemail message for Janet requesting the following:</p> <ul style="list-style-type: none"> a) Provide room number and diagram for area where I-125 sources will be stored. b) What is method for securing I-125 sources in storage area? c) What is sensitivity of detection instruments to be used to locate stray I-125 sources? We suggest licensee consider use of a thin crystal NaI probe. d) For Dr. Kagan, provide documentation that his 500 hours in brachytherapy training included the tasks listed in 35.490(b)(1)(ii). <p>On 8-31-05, Janet left a voicemail message for me stating that she received my message and forwarded it to physicist Gary Jones. On 9-1-05, Gary called and said he received the forwarded message and is attempting to obtain this information.</p> | |
| Action Required: | Await response from licensee | |
| Document Availability: | <input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date | |
| Prepared & SISP Review Completed By: | / RA / Sandy Gabriel Date: 9-1-05 | |