



UNITED STATES
NUCLEAR REGULATORY COMMISSION

REGION III
2443 WARRENVILLE ROAD STE 210
LISLE, ILLINOIS 60532-4352

AUG 24 2005

Andrea D. Browne, Ph.D.
Radiation Safety Officer
Community Hospitals of Indiana, Inc.
1500 N. Ritter Avenue
Indianapolis, IN 46219

Dear Dr. Browne:

Enclosed is Amendment No. 67 to your NRC Material License No. 13-06009-01, in accordance with your request. Please note that the changes made to your license are printed in **bold** font.

Please review the enclosed document carefully. If there are any errors or questions, please notify the U.S. Nuclear Regulatory Commission, Region III office at (630) 829-9887 so that we can provide appropriate corrections and answers.

Also, it is our understanding that following the move of the blood irradiator to its new restricted area location, that you wish authorization to release, for unrestricted use, the restricted area formerly used as the location of the blood irradiator. In that regard, promptly following the move, please provide notification of completion of the move and provide information needed by us for approval of the release, including a closeout survey and a copy of the current leak test record confirming the non-leaking status of the sealed source in the irradiator, as well as a statement confirming that you have never possessed/used any leaking sealed sources in that restricted area based on routine leak test data.

You will be periodically inspected by NRC. Failure to conduct your program in accordance with NRC regulations, license conditions, and representations made in license application and supplemental correspondence with NRC will result in enforcement action against you. This could include issuance of a notice of violation, or imposition of a civil penalty, or an order suspending, modifying or revoking your license as specified in the General Statement of Policy and Procedures for NRC Enforcement Actions. Since serious consequences to employees and the public can result from failure to comply with NRC requirements, prompt and vigorous enforcement action will be taken when dealing with licensees who do not achieve the necessary meticulous attention to detail and the high standard of compliance which NRC expects of its licensees.

This letter and the enclosed license document are exempt from public disclosure in accordance with 10 CFR 2.390, because their disclosure to unauthorized individuals could present a security vulnerability."

Sincerely,

A handwritten signature in black ink, appearing to read "Loren J. Hueter". The signature is fluid and cursive, with the first name "Loren" being more prominent.

Loren J. Hueter
Materials Licensing Branch

License No. 13-06009-01
Docket No. 030-01625

Enclosure: Amendment No. 67