

MATERIALS LICENSE

Amendment No. 44

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee		In accordance with application dated January 25, 1989,	
1. Aluminum Company of America ALCOA Technical Center		3. License number 37-07653-02 is amended in its entirety to read as follows:	
2. P.O. Box 2970 Alcoa Center, Pennsylvania 15069		4. Expiration date August 31, 1994	
		5. Docket or Reference No. 030-06172	
6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license	
A. Cesium 137	A. Sealed sources	A. See Item 9.A.	
B. Nickel 63	B. Foil contained in Hewlett-Packard Model 18713-60520 or Perkin-Elmer Model 330-0119 detector cells	B. Not to exceed 15 millicuries per foil and 150 millicuries total	
C. Hydrogen 3	C. Foil contained in AID Model 510-6007 detector cells	C. Not to exceed 200 millicuries per foil and 2 curies total	
D. Promethium 147	D. Sealed sources (Amersham Model PHC.C1)	D. Not to exceed 500 millicuries per source or 1500 millicuries total	

9. Authorized use
- A. For possession and use in Kay Ray, Accuray, Ohmart, LFE, or Texas Nuclear devices which have been evaluated and approved for licensing purposes and authorized for distribution under a license issued by the Nuclear Regulatory Commission or an Agreement State.
 - B. and C. For use in gas chromatographs for sample analysis.
 - D. For use in FAG Bearing Corporation Series FH46 gauge source holder series 9850 to measure material density.

CONDITIONS

- 10. Licensed material shall be used only at ALCOA Technical Center, ALCOA Center, and ALCOA Research Laboratory, New Kensington, Pennsylvania.
- 11. A. Licensed material shall be used by, or under the supervision of, Haig G. Sakoian or Mark Jackson.
- B. The Radiation Safety Officer for this license is Mark Jackson.

OFFICIAL RECORD COPY ML 10 1/1

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number

37-07653-02

Docket or Reference number

030-06172

Amendment No. 44

(Continued)

CONDITIONS

12. Sealed sources or detector cells containing licensed material shall not be opened or sources removed from source holders or detector cells by the licensee.
13. A(1) Any sealed source(s) or detector cell(s) specified in Items 7.B. and 7.D. shall be tested for leakage and/or contamination at intervals not to exceed 6 months and the source(s) or detector cells specified in Item(s) 7.A. shall be tested for leakage and/or contamination at intervals not to exceed 3 years. Any source received from another person which is not accompanied by a certificate indicating that a test was performed within 6 months before the transfer shall not be put into use until tested.
- (2) Notwithstanding the periodic leak test required by this condition, any licensed sealed source or detector cell is exempt from such leak tests when the source or detector cell contains 100 microcuries or less of beta and/or gamma emitting material or 10 microcuries or less of alpha emitting material.
- B. In the absence of a certificate from a transferor indicating that a test has been made within six months prior to the transfer, a sealed source or detector cell received from another person shall not be put into use until tested.
- C. Sealed sources and detector cells need not be leak tested if:
- (i) The sealed source contains only tritium; or
 - (ii) The source contains only krypton-85; or
 - (iii) The sealed source or detector cell contains not more than 100 microcuries of other beta and/or gamma emitting material, or not more than 10 microcuries of alpha emitting material; or
 - (iv) The sealed source or detector cell is in storage and not being used. However, when the source or detector cell is removed from storage for use or transfer to another person, and has not been tested within the required leak test interval, it shall be tested before use or transfer. No sealed source or detector cell shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.
- D. Tests for leakage and/or contamination shall be performed by the licensee or by other persons specifically licensed by the Commission or an Agreement State to perform such services.
14. Each gauge shall be tested for the proper operation of the on-off mechanism and indicator, if any, at no longer than six-month intervals or at such longer intervals as specified by the manufacturer, not to exceed 3 years, and at the same interval as the leak test specified in Condition 13.A.

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number

37-07653-02

Docket or Reference number

030-06172

Amendment No. 44

(Continued)

CONDITIONS

- 15. Installation, initial radiation survey, relocation, or removal from service of devices containing sealed sources shall be performed by Mark Jackson or Haig G. Sakbian or by persons specifically licensed by the Commission or an Agreement State to perform such services. Maintenance and repair of devices and installation, replacement, and disposal of sealed sources shall be performed only by persons specifically licensed by the Commission or an Agreement State to perform such services.
- 16. Prior to initial use and after installation, relocation, dismantling, alignment, or any other activity involving the source or removal of the shielding, the licensee shall assure that a radiological survey is performed to determine radiation levels around, above and below the gauge with the shutter open. This survey shall be performed only by persons specifically licensed by the Commission or an Agreement State to install gauges. A record of the results of this survey shall be maintained.
- 17. The licensee shall conduct a physical inventory every 6 months to account for all sources and/or devices received and possessed under the license. Records of inventories shall be maintained for 2 years from the date of each inventory.
- 18. Detector cells containing titanium tritide foil shall only be used in conjunction with a properly operating temperature control mechanism which prevents foil temperatures from exceeding 225 degrees Centigrade.
- 19. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents including any enclosures, listed below. The Nuclear Regulatory Commission's regulations shall govern unless the statements, representations and procedures in the licensee's application and correspondence are more restrictive than the regulations.
 - A. Letter dated January 25, 1989
 - B. Letter dated March 28, 1989
 - C. Letter dated July 20, 1989

Date SEP 12 1989

For the U.S. Nuclear Regulatory Commission
 Original Signed By:
Jack Davis
 By _____
 Nuclear Materials Safety Branch
 Region I
 King of Prussia, Pennsylvania 19406