FORM AEC-374A (2-65) 10 CFR 30

U. S. ATOMIC ENERGY COMMISSION PRODUCT MATERIAL LICE.

Page of Pages

Supplementary Sheet

License Number 37-07653-02

Amendment No. 23

Aluminum Company of America Alcoa Research Laboratories P. O. Box 772 Freeport Road New Kensington, Pennsylvania 15068

In accordance with letter dated January 15, 1971, License Number 37-07653-02 is amended as follows:

Condition 13. is amended to read:

13. Byproduct material shall be used by, or under the supervision of, individuals designated by the Isotope Committee, T. B. Bonney, Chairman.

Condition 15.F. is added:

15.F. Notwithstanding Condition 15.C., the maximum interval between leak tests for the Nuclear-Chicago Model 5120 source-device combination shall be three years.

For the U.S. Atomic Energy Commission

Original Segred By Robert E. Strakman

by Materials Branch

Division of Materials Licensing Washington, D. C. 20545

Date FEB 1 1971

REG PEBIGN 1.