

## EASTON HOSPITAL QUALITY CARE.

NMSEL

August 30, 2005

U.S.N.R.C.
Region I
Medical Licensing Section
Nuclear Materials Safety Branch
Division of Radiation Safety and Safeguards
475 Allendale Road
King of Prussia, PA 19406

To whom it may concern,

03003119

Please be advised of the following change to our N.R.C. license (#37-09938-01). We request that you amend our license accordingly.

We have closed our nuclear medicine operation at the Easton Imaging Center; 2690 Kingston Road, Easton, PA. Please remove the imaging center as a location of use.

All sealed sources have been transferred to the nuclear medicine department at Easton Hospital, 250 South 21<sup>st</sup> Street, Easton, PA. A closeout survey is enclosed.

If you have any questions, please contact our Radiation Physicist, Mark Liddington at 717-291-9813.

Sincerely,

Cornelio Catena

President/CEO

SEP/ PII

-----

Enclosure

137622

## EASTON IMAGING CENTER CLOSEOUT SURVEY OF IMAGING CENTER HOT LAB

PERFORMED BY: WALTER L. ROBINSON & ASSOCIATES

## 8/30/2005

| Countertop | Sealed Source Storage | Sink       |           |
|------------|-----------------------|------------|-----------|
| 0.02 mR/hr | 0.02 mR/hr            | 0.02 mR/hr |           |
| 22 dpm     | 28 dpm                | 34 dpm     |           |
|            |                       | 14 dom     |           |
|            | Floor                 | Γ          |           |
| }          | 0.02 mR/hr            |            | <u> </u>  |
|            | 7 dpm                 | ]          | Hot Trash |
| Floor      | •                     |            | Area      |
| 0.02 mR/hr |                       | 0          | .03 mR/hr |
| 10 dpm     |                       |            | 21 dpm    |
| 1          |                       | 1          |           |
| 1          |                       |            | 11        |
|            |                       |            |           |
|            |                       |            |           |
|            |                       |            | ļ         |
| ./         |                       |            |           |
|            |                       |            |           |

Outside Door 0.02 mR/hr

\*\*\*Survey was performed using a Ludlum 14-C G.M. survey meter.

\*\*\*Wipes tested using a Ludlum 2200 scaler and well

Mark D. Liddington
Consultant Radiation Physicist
Walter L. Robinson & Associates

| This is to acknowledge the receipt 8/3/205 includes an administrative review l | and to inform you that the initial processing which   |
|--|---|
|  | P33-cy<br>missions. Your application was assigned to a<br>e that the technical review may identify additional<br>information. |
| Please provide to this office wit  | hin 30 days of your receipt of this card  |
| Branch, who will contact you sepa Your action has been assigned Ma             | action, please refer to this control number.  |
| NRC FORM 532 (RI)<br>(6-96)  | Sincerely,<br>Licensing Assistance Team Leader  |