

U.S. NUCLEAR REGULATORY COMMISSION		Date: 9/16/05
TELEPHONE CONVERSATION RECORD		Time: 1:15 p.m.
Mail Control or Report No(s).	136166	License No(s). 29-14017-01 Docket No(s). 03001980
Name of Licensee:	Kimball Medical Center	
Name of Participant(s):	Maria Gracyk, Radiation Oncology Administrator	
Telephone No.	732-363-1900, ext 15663	
Subject: (NOTE: This will be used as the Documents Title in ADAMS)	Additional information about HDR maintenance commitment	
Summary:	<p>I called Maria to follow up on a question raised during an NRC licensing audit. Licensee's HDR application stated that all maintenance and repair will be performed by the manufacturer or individuals specifically authorized by the NRC or an Agreement State. It later stated that if an individual other than the manufacturer performs maintenance and service functions, licensee will submit to NRC a copy of the training outline and training certification as provided by the HDR manufacturer.</p> <p>Maria stated that licensee's intention is for all maintenance and repair to be provided by Nucletron, the manufacturer. In the future, if licensee wishes to consider performance of maintenance and repair by anyone other than Nucletron, they will contact NRC in writing to confirm acceptability.</p>	
Action Required:	Place into ADAMS	
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date	
Prepared & SISP Review Completed By:	/ RA / S. Gabriel Date: 9/16/05	