

U.S. NUCLEAR REGULATORY COMMISSION		Date: 09/21/2005
TELEPHONE CONVERSATION RECORD		Time: 1:05 PM
Mail Control or Report No(s).	137539	License No(s). 29-30629-01 Docket No(s). 03035672
Name of Licensee:	The Heart Center of the Oranges	
Name of Participant(s):	Elaine Rovazzi, M.S., D.A.B.R.	
Telephone No.	973-322-5118	
Subject: (NOTE: This will be used as the Documents Title in ADAMS)	RAI, Leak Test Results	
Summary:	Leak test results are needed for all sealed sources which require leak testing that were transferred to the new facility. The test result must be within 6 months of the transfer.	
Action Required:	Ms. Rovazzi said that she send them into Region I.	
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date	
Prepared & SISP Review Completed By:	/ RA / Dennis Lawyer Date: 09/21/2005	