



WEST PENN ALLEGHENY HEALTH SYSTEM
September 28, 2005

320 EAST NORTH AVENUE, PITTSBURGH, PA 15212-4772
412-359-3131

Penny Lanzisera
USNRC Region 1
Division of Nuclear Materials Safety
475 Allendale Rd.
King of Prussia, PA 19406-1415

MS 16
P-6

Licence Number 37-01317-01
Docket Number 030-02981
Mail Control No. 137476

Dear Ms. Lanzisera,

This is in response to your email letter to Joseph Rizzi (jrizzi@wpahs.org), dated September 27, 2005 at 10:16 AM.

Response to Item 1 - Authorized Users will receive vendor HDR training, including a drill of emergency response procedures prior to first clinical use. In addition, a drill of emergency response procedures will be included in the annual refresher training.

Response to Item 2 - AGH personnel providing annual refresher training and subsequent training for new employees using the HDR device will have received the training outlined in Appendix 1 of the amendment request.

Response to Item 3 - the following section will be added to the Emergency Response Procedures in Appendix 3 of the amendment request.


Access control and posting in an emergency

The HDR treatment area will be evacuated and the door locked in the event the source cannot be immediately secured. In addition, the door will be posted with "do Not Enter" or "Authorized Entry Only" signs until the source is secured.

Also, an updated list of names and telephone numbers of authorized users and authorized medical physicists will be located at the control console.

Should you have any additional questions please contact Joseph Rizzi at 412-359-0897 or Joseph Och at 412-359-6864

Sincerely,


Joseph Och, M.S., C.M.P.
Radiation Safety Officer

137476

NMSS/ROSI MATERIALS-032