

U.S. NUCLEAR REGULATORY COMMISSION		Date: September 19, 2005
TELEPHONE CONVERSATION RECORD		Time: 1:30 p.m.
Mail Control or Report No(s).	137350	License No(s). 29-03845-01 Docket No(s). 03002474
Name of Licensee:	The Valley Hospital	
Name of Participant(s):	Ki Chak, physicist	
Telephone No.	201-634-5414	
Subject: (NOTE: This will be used as the Documents Title in ADAMS)	Request to amend NRC license to remove depleted uranium shielding	
Summary:	In response to my e-mailed request, Ki said that Accelinear removed a Varian Clinac 1800. I informed Ki that Clinac 1800 linear accelerators do not contain depleted uranium (DU) shielding. I asked him to research whether The Valley Hospital previously possessed any other linear accelerators with DU shielding, such as a Clinac 4. Licensee should then submit an amendment request including information about the disposition of this linear accelerator and DU shielding.	
Action Required:	In cover letter to licensee, indicate that linac removed by Accelinear did not contain DU shielding. Await future amendment request from licensee to address DU shielding.	
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date	
Prepared & SISP Review Completed By:	/ RA / S. Gabriel	Date: 9/27/05

*e-mailed to
LAT to
place in ADAMS*