## Date: August 24, 2005 U.S. NUCLEAR REGULATORY COMMISSION Time: 2:30 PM TELEPHONE CONVERSATION RECORD Mail Control 137091 License No(s). 29-28341-02MD Docket No(s). 03031972 or Report No(s). Name of Licensee: Medi Physics, Inc. dba GE Healthcare Name of Participant(s): Richard Hughes Telephone No. 609-514-6647 or 609-209-6832 (cell) Subject: Request for Additional Information (NOTE: This will be used as the Documents Title in ADAMS) Requested additional information from the licensee's representative. Summary: 1) Iodine compounding was not described in the license renewal application. Mr. Hughes stated that licensee did not need authorization for iodine compounding at this location. 2) Licensee did not list 4 ANP's on the license renewal that were listed on the last amendment of the license. Mr. Hughes stated that two of the ANP's should be listed on the renewed license (Walter Adichie and Daryl Moyer). 3) Handling of used generators was authorized on the previous license, but was not requested on the license renewal. Mr. Hughes stated that the licensee did not need authorization for the handling of used generators. 4) The facility description in the license renewal application did not describe all areas including details for waste storage areas, locations of air hoods, installed shielding, and iodine handling facilities. Mr. Hughes stated that he would send a new floor plan with details included. 5) The licensee previously had authorization for leak testing services for customers. Mr. Hughes stated that he would check if the licensee wanted to get the authorization for this service. If so, he will use NUREG-1556, Volume 18 for guidance to provide further information. Mr. Hughes called back on 8/31/05 and stated that the licensee did not need this authorization. Action Required: Await updated floor plan and details on leak testing. Document Availability: Publicly Available Non-Publicly Available Immediate Release Normal Release Date Delay Release Date Prepared & SISP Review Completed By: /RA/ Joseph Nick Date: 8/24/05