

U.S. NUCLEAR REGULATORY COMMISSION	Date: September 19, 2005
TELEPHONE CONVERSATION RECORD	Time: 1:30 p.m.
Mail Control 137350 License No(s). 29-03845-01	Docket No(s). 03002474
Name of Licensee: The Valley Hospital Name of Participant(s): Ki Chak, physicist Telephone No. 201-634-5414 Subject: Request to amend NRC license to remove depleted uranium shielding <small>(NOTE: This will be used as the Documents Title in ADAMS)</small>	
Summary: In response to my e-mailed request, Ki said that Acceliner removed a Varian Clinac 1800. I informed Ki that Clinac 1800 linear accelerators do not contain depleted uranium (DU) shielding. I asked him to research whether The Valley Hospital previously possessed any other linear accelerators with DU shielding, such as a Clinac 4. Licensee should then submit an amendment request including information about the disposition of this linear accelerator and DU shielding.	
Action Required: In cover letter to licensee, indicate that linac removed by Acceliner did not contain DU shielding. Await future amendment request from licensee to address DU shielding.	
Document Availability: <input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date Prepared & SISP Review Completed By: / RA / S. Gabriel Date: 9/27/05	