

U.S. NUCLEAR REGULATORY COMMISSION		Date: 9/27/2005
TELEPHONE CONVERSATION RECORD		Time: 10:45
Mail Control or Report No(s).	137647	License No(s). 06-30693-01 Docket No(s). 030-35868
Name of Licensee:	Protometrix - An Invitrogen Company	
Name of Participant(s):	Gregory Waterbury, Facilities Manager and Safety Officer Steven Courtemanche, Health Physicist NRC	
Telephone No.	203-848-1104 610-337-5075	
Subject: (NOTE: This will be used as the Documents Title in ADAMS)	Telephone Deficiency	
Summary:	I spoke with Mr. Waterbury about their recent license amendment request and the additional information required. (1) I asked if they only possessed GM survey instruments. He stated that they do. I informed him that a GM survey instrument would not be the best instrument for measuring Iodine-125 contamination. I would ask them to confirm that they would obtain an instrument appropriate for the measurement of Iodine-125. (2) Personnel need to be trained in the hazards peculiar to Iodine-125 since they are different from what is currently authorized by the licensed program. (3) Recommend that the licensee perform a base-line thyroid count on individuals in case of a future contamination incident. Mr. Waterbury indicated that he would look into the requests and would like a copy of the deficiencies in writing.	
Action Required:	Place into ADAMS	
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date	
Prepared & SISP Review Completed By:	/ RA / Steven R. Courtemanche Date: 9/27/2005	