

**U.S. NUCLEAR REGULATORY COMMISSION**

Date: 9/29/05

**TELEPHONE CONVERSATION RECORD**

Time:

Mail Control 137511 License No(s). 29-17610-01 Docket No(s). 030-13049  
or Report No(s).

Name of Licensee: Virtua Memorial Hospital of Burlington County

Name of Participant(s): Paula Visconti, Ph.D. - RSO

Telephone No. 609-261-7074

Subject: RAI dated 9/29/05

(NOTE: This will be used as the  
Documents Title in ADAMS)

Summary: In the letter dated August 10, 2005, the licensee requested two physicians be added to the license for the use of 35.100; 35.200; and 35.300 materials.

1. **Sameet Rao, M.D.** - A copy of Dr. Rao's ABR certificate was provided to qualify him for the use of 35.100 and 35.200 materials. An NRC Form 313A was provided to document his training and experience with the use of 35.300 materials. However, the 313A stated that he had 25 cases of I-131 therapeutic use and did not specify whether or not the uses were for greater than or less than 33 millicurie uses of I-131. 10 CFR 390 states that the physician must show at least 3 cases of less than 33 millicuries and 3 cases of greater than 33 millicuries use of I-131. Dr. Visconti stated that he would like to get this amendment ASAP and therefore would like to add Dr. Rao for 35.100 and 35.200 only, at this time.
2. **Michael C. Brodsky, M.D.** - A copy of Dr. Brodsky's ABR certificate was provided to qualify him for the use of 35.100 and 35.200 materials. A NRC Form 313M Supplement A was provided to document his training and experience with the use of 35.300 materials. However, the document stated that he had 6 cases of hyperthyroidism and 5 cases of thyroid carcinoma. 10 CFR 35.930 states that the physician must show 10 cases of hyperthyroidism treatment. Dr. Visconti stated that he would like to get this amendment ASAP and therefore would like to add Dr. Brodsky for 35.100 and 35.200 only, at this time.
3. **Paula Visconti, Ph.D.** - A copy of her board certificate will be faxed.

Action Required: Wait for Dr. Visconti's board certificate

Document Availability:  Publicly Available  Non-Publicly Available  
 Non-Sensitive  Non-Sensitive Copyright  Sensitive  Sensitive Copyright  
 Immediate Release  Normal Release  Delay Release Date  
Prepared & SISP Review Completed By: **/ RA /** TLWeidner Date: 9/29/05