

From: Richard McKinley
To: Bsmith@prapa.com
Date: 10/20/05 1:27PM
Subject: Reference:

Reference:
CentraState Healthcare System
license 29-14966-01
docket 03008340
mail control 137603

To: Robert M. Smith

Please provide the following additional information within 30 days. You may fax your signed response to 610-337-5269, referencing mail control 137603. You may wish to leave a voice mail or e-mail message to alert me when you send me the fax.

Please send an e-mail to confirm receipt of this message.

- 1) Please describe how the RSO's training will include emergency response procedures.
- 2) Please describe your specific method for conducting daily checks of timer accuracy, including the acceptance criteria.
- 3) Please describe your specific method for conducting daily checks of the permanent area radiation monitor.
- 4) Please update your emergency procedures to include the name(s) and phone number(s) of authorized users, per 10 CFR 35.610(a)(4)(iii).

Thank you for your help. If you have any questions, you may e-mail me or call at 610-337-5102.

Richard McKinley
Health Physicist
Medical Branch
NRC Region I

CC: rwm1

Mail Envelope Properties (4357D451.C7F : 16 : 27049)

Subject: CentraState Healthcare System, 29-14966-01, 03008340, Control No.
137603

Creation Date: Thu, Oct 20, 2005 1:30 PM

From: Richard McKinley

Created By: RWM1@nrc.gov

Recipients

nrc.gov
kp1_po.KP_DO
LAT (LAT)

Post Office

kp1_po.KP_DO

Route

nrc.gov

Files

MESSAGE
Reference.mlm

Size

411
1948

Date & Time

Thursday, October 20, 2005 1:30 PM
Thursday, October 20, 2005 1:28 PM

Options

Expiration Date: None
Priority: Standard
Reply Requested: No
Return Notification: None

Concealed Subject: No
Security: Standard