

IMAGE RADIOLOGY GROUP

90 CHAMBER PLAZA, CHARLEROI, PA 15022

TOLL FREE: 1-800-222-5007 TEL: (724) 489-0263 FAX: (724) 489-0267

Licensing Assistance Section
Nuclear Materials Safety Branch
U.S. Nuclear Regulatory Commission, Region I
475 Allendale Road
King of Prussia, PA 19406-1415

K-4
MS-16

DATE:

Re: Close out of location at 5000 McKnight Road, Pittsburgh, PA 15237
License number 37-30753-01 03036077

To Whom It May Concern:

Enclosed please find the results of the close out survey and wipe tests that were conducted at the above-indicated location. The last scan that was conducted was June 2005. The close out survey was conducted on September 20, 2005.

All sealed sources that we possessed were moved to our central pharmacy for storage until the new site is ready. We have a copy of their license on file to verify that they are authorized to receive and possess these sources.

Should you require any additional information regarding this, please feel free to contact our Health Physicist, Richard W. Brown, BS, at 724-777-0144.

Sincerely,



John Teggano

enclosure

Ref: Pittsburgh Cancer Center LLC

REC'D IN LAT 10/20/05

137690
NMSS/RGNI MATERIALS-002

September 20, 2005

PET Center Close-out Survey

Image Radiology Group
5000 McKnight Road
Pittsburgh, PA 15237

NRC License #37-30753-01
PA License #PA-0885

The undersigned performed an extensive radiation contamination survey on the site previously used to house the PET Imaging center the above address. The area was surveyed with a Ludlum 14C (180753/PR085484) survey meter equipped with a pancake probe and found to have a current calibration. Background radiation was measured at 0.03 mR/hr at the time of this survey. Numerous measurements were made in areas of use and storage of licensed materials with no measurements in excess of background radiation exposure levels.

Additionally, numerous wipe tests were collected from areas of previous use and storage of licensed materials. These wipes were counted in a properly calibrated Ludlum Model 2200 well counter. No wipes were found to have activity levels above background (1,200 cpm) as recorded by this instrument on this date.

All "Caution Radioactive Materials" signs and other cautionary posting had been removed from the area at the time of this survey.

Please forward any questions pertaining to this survey to the undersigned.



Richard W. Brown, B.S.
Health Physicist
Applied Medical Physics in Radiology
724-777-0144
724-625-7520 Fax

Cardinal Health
Thomas J. Rafferty RPh.
38 Heritage Court
Delmont, Pa 15826
412-327-8429

www.cardinal.com



October 11, 2005

Pittsburgh Cancer Center
5000 McKnight Road
Pittsburgh, PA 15237

Dear John,

Cardinal Health Nuclear Pharmacy Services at 70 33rd Street Suite A in Pittsburgh, Pa has in its possession a ¹³⁷Cs Vial source S/N 16684 and a ¹³⁷Cs Rod source S/N 15736. These sources were picked up from your facility on McKnight Road, wiped tested and placed into storage at our facility. We will return these sources to you upon your direction.

Sincerely,

A handwritten signature in cursive script that reads "Jacqueline Saula".

Jacqueline Saula Pharm. D.
Radiation Safety Officer