



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

November 2, 2005

Docket No. 03036615
Control No. 137758

License No. 45-30929-01

Jane M. Coppage
Officer Manager
Prince William Nuclear Cardiology
8567-B Sudley Road
Manassas, VA 20110

SUBJECT: PRINCE WILLIAM NUCLEAR CARDIOLOGY, REQUEST FOR ADDITIONAL INFORMATION CONCERNING APPLICATION FOR AMENDMENT TO LICENSE, CONTROL NO. 137758

Dear Ms. Coppage:

This is in reference to your letter dated September 25, 2005 requesting to amend Nuclear Regulatory Commission License No. 45-30929-01. In order to continue our review, we need the following additional information:

1. Your amendment request letter should have been signed by a management representative rather than Dr. Morton E. Kalus. Please submit a letter signed by a management representative indicating that management has reviewed the application and concurs in the statements and representations contained therein. Note also that a management representative should sign all future correspondence that requests a change in your license.
2. The preceptor written attestation by Dr. Leonie Gordon did not state that Dr. Michael Escano's work experience included:
 - (A) Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;
 - (B) Performing quality control procedures on instruments used to determine the activity of dosages and performing checks for proper operation of survey meters;
 - (C) Calculating, measuring, and safely preparing patient or human research subject dosages;
 - (D) Using administrative controls to prevent a medical event involving the use of unsealed byproduct material;
 - (E) Using procedures to safely contain spilled radioactive material and using proper decontamination procedures;

(F) Administering dosages of radioactive drugs to patients or human research subjects;
and

(G) Eluting generator systems appropriate for preparation of radioactive drugs for imaging and localization studies, measuring and testing the eluate for radionuclidic purity, and processing the eluate with reagent kits to prepare labeled radioactive drugs.

These work experiences are minimum requirements as listed in 10 CFR 35.290(c)(2). Please confirm this training by an authorized preceptor written attestation.

3. The preceptor's written attestation must also state that **all** training and work experiences were satisfactorily completed. This is a requirement as stated in 10 CFR 35.290(c)(2).

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Medical, industrial, and academic uses of nuclear material**; then **toolkit index page**. Or you may obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-888-293-6498. The GPO is open from 7:00 a.m. to 9:00 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 137758. If you have any technical questions regarding this deficiency letter, please call me at (610) 337-5303.

If we do not receive a reply from you within 30 calendar days from the date of this letter, we will assume that you do not wish to pursue your application.

Sincerely,

Original signed by Thomas K. Thompson

Thomas K. Thompson
Senior Health Physicist
Commercial and R&D Branch
Division of Nuclear Materials Safety

cc:
Morton E. Kalus, M.D., Radiation Safety Officer

DOCUMENT NAME: E:\Filenet\ML053070609.wpd

SISP Review Complete: DLawyer

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI	N	DNMS/RI			
NAME	DLawyer/DRL		TThompson/TKT					
DATE	11/02/2005		11/02/2005					

OFFICIAL RECORD COPY