

**From:** Richard McKinley  
**To:** Thenzler@comcast.net  
**Date:** 11/7/05 12:42PM  
**Subject:** AMENDMENT

License No.:37-30553-01  
Docket No:03035310  
Control No: 137484

Please note that you may not reply to this letter by return e-mail. Your reply must be in writing by letter or facsimile (610-337-5269). If we do not receive a reply from you within 30 calendar days from the date of this e-mail, we will assume that you do not wish to pursue your application.

Dear Margaret:

I have a few questions regarding your emergency procedures to complete your license amendment.

- 1) If there is no patient in the room, is the source ever retracted before the vendor is called? If so, who does it?
- 2) How long do you expect it to take before Nucletron arrives in response to an emergency? If the source is exposed what will be area dose rates, and what will you do about accelerator operations?
- 3) If the patient is in the room and the source will not retract, you state that you will remove the applicator from the patient. What will you do if the source is not in the applicator?
- 4) What emergency equipment do you keep in the labelled drawer next to the sink?
- 5) Confirm that you are aware of the recordkeeping requirements of 10 CFR 35.2643.

Thank you for your help.  
Sincerely,  
Richard McKinley  
Health Physicist  
US NRC, Region I

**CC:** RWM1

**Mail Envelope Properties** (436F938C.394 : 16 : 27049)

**Subject:** Comprehensive Cancer Care, 37-30553-01, 03035310, Control No. 137484

**Creation Date:** Mon, Nov 7, 2005 12:49 PM

**From:** Richard McKinley

**Created By:** RWM1@nrc.gov

**Recipients**

nrc.gov  
kp1\_po.KP\_DO  
LAT (LAT)

**Post Office**  
kp1\_po.KP\_DO

**Route**  
nrc.gov

**Files**  
MESSAGE  
AMENDMENT.mlm

**Size**  
1404  
2133

**Date & Time**  
Monday, November 7, 2005 12:49 PM  
Monday, November 7, 2005 12:43 PM

**Options**

**Expiration Date:** None  
**Priority:** Standard  
**Reply Requested:** No  
**Return Notification:** None

**Concealed Subject:** No  
**Security:** Standard