

From: Richard McKinley
To: Sharon@npc-link.com
Date: 11/22/05 8:58AM
Subject: Renewal of Bluefield Regional Medical Center license

License No.:47-19142-01
Docket No:03017038
Control No: 137905

Sharon,

I cannot incorporate information from previous licenses into this renewal. Please re-submit the information as described below:

- 1.Please re-submit all diagrams, to scale, of all licensed material use areas. Indicate restricted and unrestricted areas contiguous with use areas (above and below). Include patient rooms (with room numbers) for in house therapy patients.
- 2.Please re-submit shielding calculations for the HDR, as well as a description of safety features of the room.
- 3.Please submit detailed procedures (not lists of procedures) for HDR emergency response and spot checks.

Please note that you may not reply to this letter by return e-mail. Your reply must be in writing by letter or facsimile (610-337-5269). If we do not receive a reply from you within 30 calendar days from the date of this e-mail, we will assume that you do not wish to pursue your application.

Thank you very much.

Sincerely,
Richard McKinley
Health Physicist
U.S. NRC, Region I

CC: Rwm1

Mail Envelope Properties (438326DB.103 : 16 : 27049)

Subject: Bluefield Regional Medical Center, 47-19142-01, 03017038, Control
No. 137905
Creation Date: Tue, Nov 22, 2005 9:10 AM
From: Richard McKinley
Created By: RWM1@nrc.gov

Recipients

nrc.gov
kp1_po.KP_DO
LAT (LAT)

Post Office
kp1_po.KP_DO

Route
nrc.gov

Files	Size	Date & Time
MESSAGE	1571	Tuesday, November 22, 2005 9:10 AM
Renewal of Bluefield Regional Medical Center license.mlm		4355 Tuesday,
November 22, 2005 9:05 AM		

Options

Expiration Date: None
Priority: Standard
Reply Requested: No
Return Notification: None

Concealed Subject: No
Security: Standard