

<b>U.S. NUCLEAR REGULATORY COMMISSION</b> <b>TELEPHONE CONVERSATION RECORD</b>		Conversation Date: August 30, 2005	
		Time: 11:00 pm	
Mail Control No.:	License No.:	Docket No.:	
Licensee/Applicant Participant(s): Ms. Dawn Edgerton	Organization: Certification Board of Nuclear Cardiology (CBNC)	Telephone No.: 240-631-8151	
Person(s) Calling: MSIB: Mohammad S. Saba - NMSS			
Subject: Followup on Changes to the Board Exam Description			
<b>Summary:</b>  The NRC staff discussed the issue of changes on the board exam . The outcome was that the examination content must include quality control procedures for dose calibrator and survey meters.			
Action Required/Taken: Review of CBNC's application for recognition will continue upon receipt of the reply from the certification board regarding the incorporation of the above changes to the examination content.			
Prepared By: Mohammad S. Saba		Date: August 30, 2005	

A-6