

**SOUTHWEST RESEARCH INSTITUTE
CALIBRATION LABORATORY
MEMORANDUM**

October 18, 2005

To: DON BANNON DIV20 B57

From: Institute Calibration Laboratory

Subject: Status of Calibration Supplier

Manufacturer/Model: OHIO SEMITRONICS PC5-103D

Description: AC WATT TRANSDUCER

Serial Number: 03080279

Asset Number: 010467

Work Order Number: 303065916

Date Calibrated: September 26, 2005

Supplier: CMI, CENTERVILLE GA - A2LA - 800 525-0408

Remarks:

☒ Supplier is on the Approved Suppliers List (ASL).

☐ Supplier is not on the Approved Suppliers List.

☒ Calibration is ISO 17025 accredited.

☐ Calibration is not ISO 17025 accredited.

☐ There is no known supplier to meet ISO 17025 accreditation at this time.

Please contact the Institute Calibration Laboratory, extension 5215, if you have any questions about the condition of this equipment or calibration documentation. Please do not remove or damage any of the calibration seals.

Attachment(s) 3

SOUTHWEST**RESEARCH****INSTITUTE**

6220 CULEBRA ROAD • POST OFFICE DRAWER 28510 • SAN ANTONIO, TEXAS, 78228-0510 • TEL (210) 522-5215 • FAX (210) 522-369

To: Don Bannon, Div 20, Bldg. 57

From: Walt Hill, Metrology Group Leader
Institute Calibration Laboratory

Date: Oct. 17, 2005

Subject: Out-of-tolerance Notice

The purpose of this notice is to alert you of a condition, which may have caused erroneous measurements affecting safety or the quality of products or services your organization provides. The attached as-found readings are provided for your evaluation to determine if the instrument listed below had an impact and if further action is required.

When the as-found results are near the specification limit, +/- a margin less than the measurement uncertainty, it is not possible to state in-tolerance or out-of-tolerance with a 95% level of confidence. It is the Institute Calibration Laboratory policy that the client is made aware of this situation because the end-user is taking some of the risk that the instrument listed below may not meet the end-user measurement requirements.

Your review/evaluation should be conducted in accordance with your organizational quality policy and procedural requirements. If we can be of further assistance, please contact the Calibration Laboratory at 522-5215.

Manufacturer: Ohio Semitronics **Model:** PC5-103D

Description: Ac Watt Transducer **Serial Number:** 03080279

Asset Number: 10467 **User ID Number:**

Last Calibration: 9/20/2004

Date Received for Service: Sep. 15, 2005 **Work Order Number:** 303065916

Service Requested: Scheduled calibration

Remarks: See certificate.

OUT OF TOLERANCE

Customer Information

Southwest Research Institute
6220 Culebra Rd
San Antonio, TX 78238

PO #: 576218J
Reference #: 443846
Account #: SO1043
SO #: 53277

Instrument Identification

Instrument Id: **03080279**

Noun: AC Watt Transducer
Mfr: Ohio Semitronics
Model: PC5-103D
Accuracy: $\pm 0.5\%$ of Full Scale
Expanded Measurement Uncertainty at K=2: $\pm 0.1\%$ of Full Scale

Serial #: 03080279

Certification Information

Reason For Service: Calibration
Type Of Calibration: Normal
As Found Condition: Out of Tolerance
As Left Condition: In Tolerance, Adjusted
Procedure: Mfr Manual :

Technician: Nym Kufuo
Cal Date: 26 SEP 05
Cal Due: 26 SEP 06
Temperature: 22.0 °C
Humidity: 44.2 %

Technician Remarks: All documented measurements are the voltmeter standard displays with the power standard outputs at the nominal test point values.
Test leads not submitted with instrument.

Quality Remarks: Calibration complies with ISO/IEC 17025 & ANSI/NCSL Z540 requirements.


This instrument has been calibrated using standards with accuracies traceable to the National Institute of Standards and Technology, derived from natural physical constants, derived from ratio measurements, or compared to consensus standards.

Reported uncertainties and/or "test uncertainty ratios" (TUR's) are expressed as expanded uncertainty values at approximately the 95% confidence level using a coverage factor of K=2. A TUR of 4:1 is routinely observed unless otherwise noted on the Certificate. Statements of compliance, where applicable, are based on test results falling within specified limits with no reduction by the uncertainty of the measurement.


Certified Measurements, Inc.'s Quality System complies with the applicable requirements of ANSI/NCSL Z540-1 and ISO/IEC 17025.

The results contained herein relate only to the item calibrated. Calibration due dates appearing on the Certificate of Calibration and label are determined by the client for administrative purposes and do not imply continued conformance to specification.

This certificate shall not be reproduced except in full, without the written permission of Certified Measurements, Inc.



Deborah J. Nazimiec, Technical Manager



Robert S. Lamb, President

Your calibration records are available on-line at <http://www.cmi-labs.com>

Printed: 29 Sep 2005 06:12 PM

510 North Houston Lake Blvd.
Centerville, Georgia 31028
Tel: 478.953.5171
Fax: 478.953.2688

CERTIFIED MEASUREMENTS, INC.
Raising the Calibration Standard...

1395 South Marietta Parkway
Building 200, Suite 206
Marietta, Georgia 30067
Tel: 770.428.1901
Fax: 770.428.7331

✓ In Tolerance ✗ Out of Tolerance

Calibration Data

Range	Nominal	As Found		As Left		Min	Max
Watt transducer, output in volts dc							
20 W	2.00	1.961	✓	2.005	✓	1.95	2.05
40 W	4.00	3.920	✗	4.010	✓	3.95	4.05
60 W	6.00	5.876	✗	6.011	✓	5.95	6.05
80 W	8.00	7.826	✗	8.007	✓	7.95	8.05
100 W	10.00	9.769	✗	9.995	✓	9.95	10.05

End of Datasheet

Calibration Standards

<u>NIST Traceable #</u>	<u>Instrument ID#</u>	<u>Description</u>	<u>Model</u>	<u>Calibration Date</u>	<u>Date Due</u>
1000178733	07754	Multimeter	8842A Opt 9	11 NOV 2004	30 NOV 2005
1001001412	2001	Calibrator, Multi-Product	5500A/SC300	24 FEB 2005	28 FEB 2006

Your calibration records are available on-line at <http://www.cmi-labs.com>

Printed: 29 Sep 2005 06:12 PM

510 North Houston Lake Blvd.
Centerville, Georgia 31028
Tel: 478.953.5171
Fax: 478.953.2688

CERTIFIED MEASUREMENTS, INC.
Raising the Calibration Standard...

1395 South Marietta Parkway
Building 200, Suite 206
Marietta, Georgia 30067
Tel: 770.428.1901
Fax: 770.428.7331