NRC FORM 591M PAR	RT 1	· <del>-</del>	· ·	U.S. NUCLEAR REGULATO	ORY COMMISSION			
(10-2003) 10 CFR 2-201 SAFET	Y INSPECTION	REPORT AND C	OMPLIANCE IN	SPECTION	Jul			
1 LICENSEE/LOCATION		1	2 NRC/REGIONAL OFFICE US NRC Kea 111					
Illiantopolis	Radiation, On	co 103 y	2443 Warranite Rd					
	Ave South	′	Sm. 4910					
Minneapolis, A	In 55435		Lisb, I) 60532					
3 DOCKET NUMBER(S)		4. LICENSEE NUMBER(S)	,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,	5. DATE(S) OF IN	ISPECTION			
630-36983		22-32583-0	1	Nov 16, 2005	•			
LICENSEE:								
Nuclear Regulatory Con	nmission (NRC) rules an	d regulations and the cond	itions of your license. Th	adiation safety and to complia se inspection consisted of sele ctor. The inspection findings a	ctive examinations			
1. Based on the inspection findings, no violations were identified.								
2. Previous vio	ation(s) closed.							
3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.								
	Non-Cited Violation(s) v	vas/were discussed involvin	g the following requiremen	nt(s) and Corrective Action(s):				
	,							
cited. This form	spection certain of your a is a NOTICE OF VIOLAT and Corrective Actions)	activities, as described belo FION, which may be subject	w and/or attached, were in to posting in accordance	n violation of NRC requirements with 10 CFR 19.11.	s and are being			
	Licensee's	s Statement of Correc	tive Actions for Item	1 4, above.				
hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.								
Title		ted Name		Signature	Date			
LICENSEE'S REPRESENTATIVE		·						
NRC INSPECTOR	G. Fark		1 State		Nov 16 Das			

NRC FORM 591M PART 1 (10-2003)

### NRC FORM 591M PART 3

(10-2003) 10 CFR 2.201

## **Docket File Information**

# SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

My

U.S. NUCLEAR REGULATORY COMMISSION

					U	
1. LICENSEE Minne REPORT NUM	apolis	Rad Onc 2005-001		NRC/REGIONAL OFFICE     Region III     2443 Warrenville Road     Lisle, IL 60532		
3. DOCKET NUMBER(S) 030-36983		4. LICENSE NUMBER(S) 22-32583-01		5. DATE(S) OF INSPECTION 11/16/05		
6. INSPECTION PROCEDURES USED 87133		7. INSPECTION FOCUS A 03.01-03.07	7. INSPECTION FOCUS AREAS 03.01-03.07			
			SUPPLEMENTAL INSP	ECTION INFORMATION		
1. PROGRAM 2230	I CODE(S)	2. PRIORITY	3. LICENSEE CONTACT Mary Fox		4. TELEPHONE NUMBER 612/900-8777	
X Main Office Inspection Field Office				Next Inspection Date:	11/2007	
Те	emporary Jo	ob Site				

#### **PROGRAM SCOPE**

Licensee is a unit of a large hospital group located in Minneapolis. Minnesota Licensee is a stand alone radiation oncology unit of Fairview Southdale Hospital conducting treatments using an HDR unit. The licensee plans on performing seven procedures per week using an HDR unit. This licensee owns one HDR unit which is currently stored with the manufacturer. Licensee does perform conventional brachytherapy under the main hospital license.

### Performance Observations

The inspector toured the facilities and interviewed authorized users/staff members. Each appeared knowledgeable in radiation safety and isotope handling techniques. Package receipt procedures were demonstrated for the inspector as well as rad waste handling practices. Independent surveys by the inspector did not detect any abnormal reading and were within the expected range.

The inspector reviewed the written directives for conventional brachytherapy under the main hospital license. No HDR treatments were reviewed as the unit is in storage at the manufacturer. No abnormalities were observed.