

NRC FORM 691M PART 1
(11-2003)
10 CFR 2.201

U.S. NUCLEAR REGULATORY COMMISSION

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Warren Radiology, PC 27075 Schoenerr Rd Warren, MI 48088 REPORT NUMBER(S) 2005/001		2. NRC/REGIONAL OFFICE REGION III US NUCLEAR REGULATORY COMMISSION 801 WARRENVILLE ROAD LISLE IL 60532-4351	
3. DOCKET NUMBER(S) 030-02154	4. LICENSE NUMBER(S) 21-13388-01	5. DATE(S) OF INSPECTION 11/15/2005	

LICENSEE:
The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the Inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the Inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.



_____ non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE	WARREN RADIOLOGY		11-18-05
NRC INSPECTOR	Sarah R. Bakhsh		11/18/2005 15 85

Docket File Information
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Warren Radiology, PC REPORT NUMBER(S) 2005/001		2. NRC/REGIONAL OFFICE REGION III	
3. DOCKET NUMBER(S) 030-02154	4. LICENSE NUMBER(S) 21-13388-01	5. DATE(S) OF INSPECTION 11/16/2005	
6. INSPECTION PROCEDURES USED 87130	7. INSPECTION FOCUS AREAS 03.01-03.07		

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02201	2. PRIORITY 5	3. LICENSEE CONTACT Del Charbonier, RSO	4. TELEPHONE NUMBER 586-758-5800
-----------------------------	------------------	--	-------------------------------------

Main Office Inspection Next Inspection Date: 11/2010

Field Office Inspection _____

Temporary Job Site Inspection _____

PROGRAM SCOPE

The licensee was a small clinic located in Warren, Michigan. The licensee was authorized to use byproduct materials under Sections 35.100 and 35.200. The licensee had one full-time and one part-time nuclear medicine technologists and performed approximately 8-12 procedures daily and used technetium-99m for stress tests, bone scans, thyroid and occasional liver and renal scans. The technetium-99m was received in unit doses and in bulk from a local pharmacy. All waste was held for decay-in-storage or returned to the radiopharmacy.

PERFORMANCE OBSERVATIONS

During the inspection, the inspector: (1) observed adequate security of the licensed material; (2) observed adequate posting of areas; (3) observed nuclear medicine technologist perform survey meter QC; (4) observed package receipt and return surveys; (5) observed daily and weekly contamination surveys; (6) reviewed select records; and (7) observed demonstrations of emergency response procedures.