

Vision Imaging of Lebanon, LLC.

805 Helen Drive
Lebanon, PA 17042

December 6, 2005

Dennis Lawyer
United States Nuclear Regulatory Commission
Region 1
475 Allendale Road
King of Prussia, PA 19406

J-6
MS-16

Re: Mail Control Number 137963

37-31108-01
03037097

2005 DEC 08 AM 10:33

RECEIVED
NRC

Dear Dennis,

In response to your questions:

1. We request that Larry Braunstein, M.D. whom is listed as an authorized user and is the RSO on NRC license number 37-30871-01 be listed as an authorized user and the RSO on the NRC license for Vision Imaging of Lebanon, LLC. Dr. Braunstein's credentials should be on file with the NRC.
3. Radiation monitoring equipment will be calibrated by a person qualified to perform survey meter calibration.
4. We reserve the right to upgrade our survey instruments as necessary as long as they are adequate to measure the type and level of radiation for which they are used.
5. We will calibrate equipment used to measure dosages with nationally recognized standards or manufacturers' instructions.
6. We will monitor external and/or internal occupational exposure by providing dosimetry.
7. We will develop, implement and maintain written procedures for area surveys in accordance with 10CFR20.1101 that meet the requirement of 10 CFR 20.1505 and 10 CFR 35.70.
8. We will develop and implement and maintain procedures for safe use of unsealed byproduct material.

If you need anything else, please feel free to contact me.

Sincerely,



Francisco J. Morales, M.D.
President

137963

NMSS/RGNI MATERIALS-002