

U.S. NUCLEAR REGULATORY COMMISSION		Date: 12/08/2005
TELEPHONE CONVERSATION RECORD		Time: 9:05 AM
Mail Control 137845 or Report No(s).	License No(s). 07-03990-01	Docket No(s). 03003868
<p>Name of Licensee: Astrazeneca Pharmaceuticals LP</p> <p>Name of Participant(s): Scott Petlick</p> <p>Telephone No. 302-886-1083</p> <p>Subject: RAI, Additional Comments to letter dated December 1, 2005 <small>(NOTE: This will be used as the Documents Title in ADAMS)</small></p>		
<p>Summary:</p> <p style="margin-left: 40px;">I talked to Mr. Petlick about the Nickel-63 sources and that I needed to get the model numbers associated with them. He stated that they didn't have any of them any more and that they had been transferred. I then stated that he should send me the leak tests within 6 months and the transfer date associated with all of the tritium sealed sources.</p> <p style="margin-left: 40px;">We discussed the release criteria of item #10. I state that he should request the release criteria as listed in Volume 7 tables Q.2 and Q.3 as appropriate for that material. He may then state except for the value of I-125 and use the values that are in table R.3 of Volume 9 for equipment.</p>		
<p>Action Required: Mr. Petlick stated that he would get the Nickel-63 leak test and transfer data and review the release criteria of Volume 7 tables Q.2 and Q.3.</p>		
<p>Document Availability: <input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available</p> <p><input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright</p> <p><input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date</p> <p>Prepared & SISP Review Completed By: / RA / Dennis Lawyer Date: 12/08/2005</p>		