

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Genesys Hurley Cancer 302 Kensington Ave Flint, MI 48503 REPORT 2005001		2. NRC/REGIONAL OFFICE US NRC Reg III 2443 Warmville Rd Suite 210 Lisle, IL 60532	
3. DOCKET NUMBER(S) 03036106	4. LICENSEE NUMBER(S) 21-22222-01	5. DATE(S) OF INSPECTION Nov 30, 2005	

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

1. Based on the inspection findings, no violations were identified.

2. Previous violation(s) closed.

3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	G. Park		11/30/05

Docket File Information
SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION



1. LICENSEE Genesys Hurley Cancer Inst REPORT NUMBER(S) 2005-001		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road Lisle, IL 60532	
3. DOCKET NUMBER(S) 030-36106		4. LICENSE NUMBER(S) 21-32322-01	
5. DATE(S) OF INSPECTION Nov 30, 2005		6. INSPECTION PROCEDURES USED 87133	
7. INSPECTION FOCUS AREAS 03.01-03.07		SUPPLEMENTAL INSPECTION INFORMATION	
1. PROGRAM CODE(S) 2230	2. PRIORITY 2	3. LICENSEE CONTACT Ahmed Aki, M.D.	4. TELEPHONE NUMBER 810/762-8490
<input checked="" type="checkbox"/> Main Office Inspection		Next Inspection Date: 11/2007	
<input type="checkbox"/> Field Office _____			
<input type="checkbox"/> Temporary Job Site _____			

PROGRAM SCOPE

Licensee is a specialty clinic located in Flint, Michigan. Licensee's main focus is on cancer treatment using a HDR unit. The licensee performs approximately 4 procedures per day. This clinic is equipped with a nucletron HDR unit backed by a linac unit. No conventional brachytherapy is performed at this facility

Performance Observations

The inspector toured the facilities and interviewed authorized users/staff members encountered. Each appeared knowledgeable in radiation safety and in specific procedures associated with the HDR unit. Receipt of new sources was demonstrated for the inspector as well as discussions on staff training were conducted. Independent surveys by the inspector around the unit did not detect any abnormal reading and were within the expected range.

The inspector observed the staff prepare the unit and conduct the daily safety checks. An adequate knowledge of radiation safety was demonstrated.

No violations were noted.