U.S. NUCLEAR REGULATORY COMMISSION					Date:	12/5/2005	
TELEPHONE CONVERSATION RECORD				D -	Time:	13:30	
Mail Control or Report No(s).	2005-001	License No(	s). 45-25406-0	1	[	Docket No(s).	030-34544
Name of Licensee:		Virginia Cardiovascular Specialists, PC					
Name of Participant(s):		Yvonne J. Weaver, M.D., Radiation Safety Officer Steven R. Courtemanche, Health Physicist NRC					
Telephone No.		804-323-5011 610-337-5075					
Subject: (NOTE: This will be used as the Documents Title in ADAMS)		Facility Diagrams, Multiple RSO					
Summary:		I called Dr. Weaver to inform her of my review of the documents depicting the facilities and the discussion with James Dwyer, Branch Chief on the requirements concerning Radiation Safety Officers (RSO). The diagrams were accurate in that they showed all of the locations of use within their facilities. The licensee would not be required to submit new facility diagrams. Dr. Weaver wanted documentation in the regulations stating that only one RSO was permitted on a license. I informed her that a Clear Letter would be issued from the Region I Office and that a statement that the regulations indicated an individual is to be named on the license as the RSO not individuals. Also, NUREG-1556, Vol. 9 indicates that there should only be one RSO. Guidance from NRC HQ is that one RSO is to be named but that individuals can act on behalf of the RSO when multiple sites are involved. I thanked Dr. Weaver for her patience in this matter.					
Action Required:		Place in ADAMS					
Document Availability:		Х	Publicly Available	•		Non-Publicly	Available
X Non-Sensit	ive	Non-Sensiti	ve Copyright	Sens	itive	Sensitive	e Copyright
Immediate	Release	X Nor	mal Release		_ Dela	ay Release Dat	е
Prepared & SISP Review Completed By:		I RA I	Steven Courtem	anche	Date:	12/5/2005	