

U.S. NUCLEAR REGULATORY COMMISSION TELEPHONE CONVERSATION RECORD	Date: 12/15/05 Time: 9:30 am
Mail Control 137712 License No(s). 55-17986-01 or Report No(s).	Docket No(s). 03013764
Name of Licensee: Schneider Regional Medical Center	
Name of Participant(s): Charles Chipley Richard McKinley	
Telephone No. (340)513-2852 (610)337-5102	
Subject: Amendment adding HDR <small>(NOTE: This will be used as the Documents Title in ADAMS)</small>	
Summary: Mr. Chipley confirmed that each of the second 4 conditions listed in the license amendment cover letter will be met, and that his understanding is the same as that of Mr. McKinley. He explained that the fourth item stemmed from Varian's procedures which actually referred to the disposition of the source after it has been placed into the emergency container. He agreed that licensee personnel were authorized to place an unshielded source in the emergency container.	
Action Required: Issue amendment.	
Document Availability: <input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date	
Prepared & SISP Review Completed By: / RA / Richard McKinley Date: 12/15/05	